



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001 & ISO 13485 Certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : 874/L

Date: 28-11-2018

PRODUCT :Truenat™ H1N1

LOT. NO. : HN024

MFG. DATE : 2018-11

EXP. DATE : 2020-10

PACK SIZE : 5T, 20T

STORAGE : 2°C to 30°C

A.R. NO. : B-HN-024

PRECAUTION: The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01no.	Pass

Performance			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
Specificity Tests		HN	Control
Negative Control		ND	29.75
Positive Control		30	30.13
		28.33	28.67
		27	27
		29.14	29
Standard deviation		± 1.28	± 1.29

*ND- Not detected

*A deviation of ≤ 1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA

Works: Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: 91 832 6624783/6682000 **Fax:** 91 832 2783139 **E mail:** admin@molbiodiagnostics.com

Office: No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA

Tel: 91 832 2458111/2458112 **Email:** sales@molbiodiagnostics.com

Website: www.molbiodiagnostics.com **CIN:** U33125GA2000PTC002909