

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000018

Date:25/07/2022

Doc ID: COA/TN/TX075d

BRAND NAME	Truenat® MTB-RIF Dx			
GENERIC NAME	Chip-based Real Time PCR Test for Rifampicin Resistant Mycobacterium tuberculosis			
LOT No.	TX075	CHIP LOT NO.	HDLW442	
MFG. DATE	2022-03	EXP. DATE	2024-02	
PACK SIZE	5T, 20T, 25T, 50T			
STORAGE	2°C to 30°C A.R. No.:E			

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation Pass	
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no, 25 no 50 no.		
2	Colour of dessicant pouch	Orange	Pass	
3	Package insert	01 no.	Pass	

	PERFORMANCE							
Specificity	Tm							
Tests	P1	P2	Control	Result				
	66.65	67.07	70.36	RIF RESISTANCE DETECTED				
Sample 1	66.66	67.03	70.82	RIF RESISTANCE DETECTED				
	66.52	67.12	70.63	RIF RESISTANCE DETECTED				
	66.8	67.12	70.26	RIF RESISTANCE DETECTED				
	66.86	72.24	70.43	RIF RESISTANCE NOT DETECTED				
Sample 2	67	72.36	69.94	RIF RESISTANCE NOT DETECTED				
	66.96	72.33	69.78	RIF RESISTANCE NOT DETECTED				
	67.05	72.4	69.92	RIF RESISTANCE NOT DETECTED				
	67.07	67.07	69.5	RIF RESISTANCE NOT DETECTED				
	66.61	66.61	70.38	RIF RESISTANCE NOT DETECTED				

*Tm-Melting temperature, RIF-Rifampicin

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL

Name: Dr. Praveen Kumar

Designation: QC Head

Page 1 of 1