



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000018

Date: 13/07/2022

Doc ID: COA/TN/CT019

BRAND NAME	Truenat [®] CT		
GENERIC NAME	Chip-based Real Time PCR Test for <i>Chlamydia trachomatis</i>		
LOT No.	CT019	CHIP LOT NO.	HDLW436
MFG. DATE	2021-11	EXP. DATE	2023-10
PACK SIZE	5T, 20T, 50T		
STORAGE	2°C to 30°C		A.R. No. : B-CT-019a

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.


Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 no, 20 no, 50 no.	Pass
2	Colour of desiccant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE		
CONTROLS	(Ct Values)	
	CLT	Control
Specificity Tests		
Negative Control	ND	27.5
Positive Control	28.88	24.14
	28.75	23.6
	28.67	23.75
	28.2	24.13
Average	28.63	23.91
Standard deviation	± 0.30	± 0.27

*ND- Not detected, Ct- Cycle threshold, CLT- *Chlamydia trachomatis*

*A deviation of ≤ 1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head
Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO