

MOLBIO DIAGNOSTICS PVT. LTD. An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 18/03/2020

Product: Trueprep® AUTO Universal Cartridge Based Sample prep kit

Trueprep® AUTO Universal

Pack size: 5, 25 TESTS

Lot No.: USPK280 Mfg Date: 2020-02 Exp Date: 2022-01

Cartridge Based

A.R. No.: B-USPK-280

Sample prep kit

Storage: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab™ micro PCR Analyzers						
Sr. No.	Sample Negative/Positive	Negative: Test- ND Positive: Test- Detected	IC- Detected IC- Detected			
		Test (Ct)	C (Ct)			
1	Negative MTB	ND	28.8			
2	Positive MTB	31.29	28.5			
3	Negative HBV	ND	30.33			
4	Positive HBV	29.14	28.5			
5	Positive HIV	29.2	28.71			
6	Positive HCV	30.21	29.0			

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SAME OF STREET

QUALLOX CONTROL

Name: Dr. Prayden Lumar

Designation: QC Head

^{*}IC- Internal Control
*ND- Not Detected