

MOLBIO DIAGNOSTICS PVT. LTD. An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 19/03/2020

PRODUCT: Trueprep™ AUTO MTB Sample Pre-treatment Pack

Trueprep™ AUTO MTB

Lot No.: MSPT049 Mfg Date: 2019-08 Exp Date: 2021-07

Sample Pre-treatment Pack

A.R. No.: B-MSPT-049 STORAGE: 2°C to 40°C Pack size: 20 TESTS, 5 TESTS

PRECAUTION: Refer Safety Data Sheet and follow instructions for use

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (MTB Sample Pre- treatment Pack)	(Present (P)/Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Liquefaction Buffer	Present	Present	
3	Disposable Transfer Pipette (1ml)	Present	Present	

	Performa	nce on Tr	uelab™ Real Time micro PCR A	nalyzers	
Sr. No.	Controls	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Neg	ative Control				
1	Sample 1	ND	Control should be detected	ND	29.33
Posi	tive Control				
1	Sample 1	Test a	nd Control should be detected	31	28.17
2	Sample 2	Test a	nd Control should be detected	29.71	27.43
3	Sample 3	Test a	and Control should be detected	31	27.6

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURING PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED

> QUALITY CONTROL Name: Dr. Praveen Kumar Designation: QC Head

Page 1 of 1

^{*}IC- Internal Control

^{*}ND- Not Detected