



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001:2015 & ISO 13485:2016 Certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : MFG/IVD/2019/000018

Date: 07-04-2020

BRAND NAME	Truenat™ Beta Cov
GENERIC NAME	Chip-based Real Time PCR Test for Beta Coronavirus
LOT No.	CV001
MFG. DATE	2020-04
EXP. DATE	2022-03
PACK SIZE	5T, 20T
STORAGE	2°C to 30°C
A.R. No.:	B-CV-001

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 no, 20 no.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

PERFORMANCE	
CONTROLS (Ct Values)	
Specificity Tests	COV Control
Negative Control	ND
Positive Control	ND
	25.75
	31.33
	26.33
	27.2
	31.5
	27.33
	30.0
	26.8
	31.17
	27.33
	30.75
	27.5
	30.43
	26.29
	31.14
	26.6
Standard deviation	+ 0.53
	± 0.47

*ND- Not detected, Ct- Cycle threshold, COV- Coronavirus

*A deviation of ≤ 1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOTS QC APPROVED FOR SALE

QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head
Page 1 of 1

Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722, INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnosics.com

Email: sales@molbiodiagnosics.com Website: www.molbiodiagnosics.com

CIN: U33125GA2000PTC002909 GSTN: 30AACCM8275E1Z0