



MOLBIO DIAGNOSTICS PVT. LTD.
An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 24/03/2020

Product: Trueprep[®] AUTO Universal Cartridge Based Sample prep kit

Trueprep[®] AUTO Universal
Cartridge Based
Sample prep kit

Lot No.: USPK282 Mfg Date: 2020-02 Exp Date: 2022-01

A.R. No.: B-USPK-282

Pack size: 5, 25 TESTS

Storage: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep [®] AUTO Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab[™] micro PCR Analyzers

Sr. No.	Sample Negative/Positive	Negative: Test- ND IC- Detected Positive: Test- Detected IC- Detected	
		Test (Ct)	IC (Ct)
1	Negative MTB	ND	27.8
2	Positive MTB	30	28.33
3	Positive MTB	31.67	29.8
4	Positive MTB	31.67	29.75

*Ct- Cycle threshold

*IC- Internal Control

*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



QUALITY CONTROL

Name: Dr. Praveen Kumar

Designation: QC Head

Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0