



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : MFG/IVD/2019/000018

Date:18/10/2022
Doc ID: COA/TN/ML019c

BRAND NAME	Truenat® Malaria Pv/Pf		
GENERIC NAME	Chip-based Real Time Duplex PCR Test for Plasmodium vivax and Plasmodium falciparum		
LOT No.	ML019	CHIP LOT NO.	MGHW011
MFG. DATE	2022-09	EXP. DATE	2024-08
PACK SIZE	5T, 20T, 25T, 50T		
STORAGE	2°C to 30°C		A.R. No.:B-ML-019c

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no,25 no. 50 nos.	Pass
2	Colour of desiccant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE			
CONTROLS	Ct Values		
	Pv	Pf	Control
Specificity Tests			
Negative Control	ND	ND	26.33
Positive Control (Pv)	32.4	ND	25.67
Positive Control (Pf)	ND	31.13	25.2
Positive Control (Pv and Pf)	30.43	29.36	25.33
	30.57	29.83	26
Average	30.5	29.6	26
Standard deviation	+ 0.10	+ 0.33	+ 0.47

*ND- Not detected, Ct- Cycle threshold, Pv- Plasmodium vivax, Pf- Plasmodium falciparum

*A deviation of ≤ 1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE

QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head
Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO