



MOLBIO DIAGNOSTICS PVT. LTD.  
An ISO 9001:2015 & ISO 13485:2016 certified company

### CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 06/04/2020

Product: Trueprep <sup>®</sup> AUTO Universal Cartridge Based Sample prep kit	
Trueprep <sup>®</sup> AUTO Universal Cartridge Based Sample prep kit	Lot No.: USPK284 Mfg Date: 2020-02 Exp Date: 2022-01
	A.R. No.: B-USPK-284
Pack size: 5, 25 TESTS	Storage: 2°C to 40°C

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep <sup>®</sup> AUTO Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab <sup>™</sup> micro PCR Analyzers			
Sr. No.	Sample Negative/Positive	Negative: Test- ND	IC- Detected
		Positive: Test- Detected	IC- Detected
		Test (Ct)	IC (Ct)
1	Negative MTB	ND	27.14
2	Positive MTB	26.6	26.83
3	Positive MTB	26	26.8
4	Positive MTB	26.25	26.8

\*Ct- Cycle threshold

\*IC- Internal Control

\*ND- Not Detected

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE



**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

## Molbio Diagnostics Pvt. Ltd.

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0