

MOLBIO DIAGNOSTICS PVT. LTD. An ISO 9001:2015 & ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg.Lic. No.: MFG/IVD/2019/000003

Date: 14/04/2020

PRODUCT: Trueprep® AUTO Universal Sample Pre-treatment Pack

Trueprep® AUTO Universal

Lot No.: USPT035 Mfg Date: 2020-03 Exp Date: 2022-02

Sample Pre-treatment Pack A.R. No.: B-USPT-035

Pack size: 20 TESTS, 5 TESTS

STORAGE: 2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
A	Contents in the carton (Trueprep® AUTO Universal Sample Pre- treatment Pack)(Present (P)/Not Present (NP))	(Present (P)/ Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Disposable Transfer Pipette (1ml)	Present	Present	

	Perform	ance on	Truelab™ micro PCR Analyzer	S	
Sr. No.	Control	Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test	IC
Negat	tive Control				
1	Neg HBV	ND	Control should be detected	ND	29.75
Positi	ve Control				
1	Pos HBV	Test and Control should be detected		29.13	29.6
2	Pos HIV	Test and Control should be detected		29	29.67
3	Pos HCV	Test and Control should be detected		30.67	30.6

^{*}Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL Name: Dr. Praveen Kumar Designation:QC Head

Page 1 of 1

^{*}IC- Internal Control

^{*}ND- Not Detected