



**MOLBIO DIAGNOSTICS PVT. LTD.**  
An ISO 9001 & ISO 13485 certified company  
**CERTIFICATE OF ANALYSIS**

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 29-04-2020

PRODUCT : Truenat™ Universal Control Kit  
LOT. NO. : UC020  
MFG. DATE : 2019-12  
EXP. DATE : 2020-11  
PACK SIZE : 08 T  
STORAGE : 2°C to 40°C

A.R. No. : B-UC-020

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Negative control	Present	Present
2	Positive control (08- strip tubes containing dried down positive control)	Present	Present

Specification (Ct Value)				Observation (Ct Value)	
Sr. No.	Test	Positive control	Negative control	Positive control	Negative control
01	MTB	29.08 ± 2	ND	29.08	ND
02	MTB Plus	27.17 ± 2	ND	27.17	ND
03	Beta CoV	28.36 ± 2	ND	28.36	ND
04	MTB-RIF Dx	RIF resistance not detected	Indeterminate	RIF resistance not detected	Indeterminate

\*ND- Not detected, Ct- Cycle threshold, RIF- Rifampicin

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**

  
**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

**Molbio Diagnostics Pvt. Ltd.**

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0