



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS OF FINISHED PRODUCT

Mfg. Lic. No.: MFG/IVD/2019/000018

Date:09/12/2022  
Doc ID: COA/TN/TX080R/d

66. BRAND NAME	Truenat <sup>®</sup> MTB-RIF Dx		
GENERIC NAME	Chip-based Real Time PCR Test for Rifampicin Resistant <i>Mycobacterium tuberculosis</i>		
LOT No.	TX080 R	CHIP LOT NO.	HDLW516
MFG. DATE	2022-08	EXP. DATE	2024-07
PACK SIZE	5T, 20T, 25T, 50T		
STORAGE	2°C to 30°C		A.R. No.:B-TX-080R/d

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.


Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no, 25 no 50 no.	Pass
2	Colour of dessicant pouch	Orange	Pass
3	Package insert	01 no.	Pass

PERFORMANCE

Specificity Tests	Tm		Control	Result
	P1	P2		
Sample 1	66.77	67.21	69.92	RIF RESISTANCE DETECTED
	66.7	66.86	70.06	RIF RESISTANCE DETECTED
	66.71	66.71	70.26	RIF RESISTANCE DETECTED
	65.29	65.4	71.08	RIF RESISTANCE DETECTED
Sample 2	66.47	72.14	69.99	RIF RESISTANCE NOT DETECTED
	66.1	71.52	70.57	RIF RESISTANCE NOT DETECTED
	66.89	72.26	69.8	RIF RESISTANCE NOT DETECTED
	66.4	72.4	70.22	RIF RESISTANCE NOT DETECTED
	66.85	72.5	69.45	RIF RESISTANCE NOT DETECTED
	67.05	72.14	69.62	RIF RESISTANCE NOT DETECTED

\*Tm-Melting temperature, RIF-Rifampicin

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

  
QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head  
Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0