



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License. No.: MFG/IVD/2019/000003

Date:19/01/2023
Doc ID:COA/TP/MSPT165

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack
Generic Name	Sample pre-treatment pack for MTB samples
Lot No.	MSPT165
Mfg. Date	2023-01
Exp. Date	2024-12
Pack Size	50T, 25T, 20T, 05T
Storage condition	2°C to 40°C

PREAUTION: Refer Safety Data Sheet and follow instructions for use.

Physical verification data			
Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1ml- graduated)	Present	Present
4	Package Insert	Present	Present

FM/QT/224-00

Page 1 of 2

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO



MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License. No.: MFG/IVD/2019/000003

Date:19/01/2023
Doc ID:COA/TP/MSPT165

Performance testing data					
Performance on Truelab® Real Time micro PCR Analyzers					
Sr. No.	Controls	Specification (Ct Values)		Test Lot (Ct Values) Average	
1	Specificity Tests	Test	IC	Test	IC
Negative Control					
1	Negative MTB	ND	Control should be detected	ND	27.59
2	Negative MTB	ND	Control should be detected	ND	27.11
3	Negative MTB	ND	Control should be detected	ND	27.7
Positive Control					
1	Positive MTB	Test and Control should be detected		28.46	27.68
2	Positive MTB	Test and Control should be detected		28.63	27.48
3	Positive MTB	Test and Control should be detected		28.22	27.43

*Ct- Cycle threshold *IC- Internal Control *ND- Not Detected

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

FM/QT/224-00

Page 2 of 2

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO