

## MOLBIO DIAGNOSTICS PRIVATE LIMITED

## ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 28/02/2023

Doc ID:COA/TP/MSPT169

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack	
Generic Name	Sample pre-treament pack for MTB samples	
Lot No.	MSPT169	
Mfg. Date	2023-02	
Exp. Date	2025-01	
Pack Size	5T,20T,25T, 50T	
Storage condition	2°C to 40°C	

PRECAUTION: Refer Safety Data Sheet and follow instructions for use

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present Present	
4	Package Insert	Present	Present

FM/QT/224-00



	Performand	e on Truelab	o <sup>®</sup> Real Time micro P	CR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
1	Specificity Tests	Test	IC	Test 1 (Average)	IC (Average)
Negative (	Control				
1	Negative MTB	ND	Control should be detected	ND	27.80
2	Negative MTB	ND		ND	27.86
3	Negative MTB	ND		ND	27.58
Positive C	ontrol		-h		
1	Positive MTB	Test and Control should be detected		29.31	27.53
2	Positive MTB	Test and Control should be detected		29.72	27.64
3	Positive MTB	Test and Control should be detected		29.31	27.83

\*Ct- Cycle Threshold

\*IC- Internal Control \*ND- Not Detected \*NA- Not Applicable

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

> Name: Dr. Praveen Kumar MK **Department: Quality Control** Designation: QC Head