

MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003 Date: 01/03/2023

Doc ID:COA/TP/MSPT170

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack			
Generic Name	Sample pre-treament pack for MTB samples			
Lot No.	MSPT170			
Mfg. Date	2023-03			
Exp. Date	2025-02			
Pack Size	5T,20T,25T, 50T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present



	Performand	e on Truelab	[®] Real Time micro P	CR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test 1 (Average)	IC (Average)
Negative (Control				
1	Negative MTB	ND	Control should be detected	ND	27.70
2	Negative MTB	ND		ND	27.74
3	Negative MTB	ND		ND	27.42
Positive C	ontrol				
1	Positive MTB	Test and Control should be detected		28.43	26.83
2	Positive MTB	Test and Control should be detected		29.65	27.00
3	Positive MTB	Test and Control should be detected		29.46	26.80

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name: Dr.Praveen Kumar MK Department: Quality Control Designation: QC Head