

## MOLBIO DIAGNOSTICS PRIVATE LIMITED

## ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 11/03/2023

Doc ID:COA/TP/MSPT176

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack		
Generic Name	Sample pre-treament pack for MTB samples		
Lot No.	MSPT176		
Mfg. Date	2023-03		
Exp. Date	2025-02		
Pack Size	5T,20T,25T, 50T		
Storage condition	2°C to 40°C		

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation	
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))		
1	Lysis Buffer	Present	Present	
2	Liquefaction Buffer	Present	Present	
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present	
4	Package Insert	Present	Present	

FM/QT/224-00

Page 1 of 2



	Performand	e on Truela	b <sup>®</sup> Real Time micro P	CR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
		Neg	gative Control		
1	Negative MTB	ND	Control should be detected	ND	27.22
2	Negative MTB	ND	Control should be detected	ND	28.62
3	Negative MTB	ND	Control should be detected	ND	27.55
		Pos	sitive Control		
1	Positive MTB	Test and Control should be detected		29.99	27.84
2	Positive MTB	Test and Control should be detected		29.10	27.72
3	Positive MTB	Test and Control should be detected		28.63	27.41

<sup>\*</sup>Ct- Cycle Threshold

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose. NOS

> Name: Dr. Praveen Kumar MK **Department:** Quality Control Designation: QC Head

<sup>\*</sup>IC- Internal Control \*ND- Not Detected \*NA- Not Applicable