

## **MOLBIO DIAGNOSTICS PRIVATE LIMITED**

## ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003 Date: 15/03/2023

Doc ID:COA/TP/MSPT178

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack		
Generic Name	Sample pre-treament pack for MTB samples		
Lot No.	MSPT178		
Mfg. Date	2023-03		
Exp. Date	2025-02		
Pack Size	5T,20T,25T, 50T		
Storage condition	2°C to 40°C		

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present



	Performanc	e on Truelal	b <sup>®</sup> Real Time micro P	CR Analyzer		
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)	
1	Specificity Tests	Test	IC	Test (Average)	IC (Average)	
		Neg	ative Control			
1	Negative MTB	ND	Control should be detected	ND	28.98	
		Pos	sitive Control			
1	Positive MTB	Test and Control should be detected		29.33	27.30	
2	Positive MTB	Test and Control should be detected		30.20	27.99	

\*Ct- Cycle Threshold

\*IC- Internal Control

\*ND- Not Detected

\*NA- Not Applicable

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name: Dr.Praveen Kumar MK
Department: Quality Control
Designation: QC Head