

MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 15/03/2023

Doc ID:COA/TP/MSPT179

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack		
Generic Name	Sample pre-treament pack for MTB samples		
Lot No.	MSPT179		
Mfg. Date	2023-03		
Exp. Date	2025-02		
Pack Size	5T,20T,25T, 50T		
Storage condition	2°C to 40°C		

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

FM/QT/224-00



	Performanc	e on Truelal	b® Real Time micro F	CR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
		Neg	ative Control		
1	Negative MTB	ND	Control should be detected	ND	29.00
		Pos	sitive Control		
1	Positive MTB	Test and Control should be detected		29.91	27.55
2	Positive MTB	Test and Control should be detected		30.24	27.35

*Ct- Cycle Threshold

*IC- Internal Control *ND- Not Detected

*NA- Not Applicable

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

> Name: Dr.Praveen Kumar MK **Department: Quality Control** Designation: QC Head