

MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 07/03/2023

Doc ID:COA/TP/MSPT172

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack	
Generic Name	Sample pre-treament pack for MTB samples	
Lot No.	MSPT172	
Mfg. Date	2023-03	
Exp. Date	2025-02	
Pack Size	5T,20T,25T, 50T	
Storage condition	2°C to 40°C	

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present



	Performand	e on Truelat	o [®] Real Time micro P	CR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative (Control				
1	Negative MTB	ND	Control should be detected	ND	28.41
2	Negative MTB	ND		ND	28.32
3	Negative MTB	ND		ND	28.41
Positive C	ontrol		-ll-		
1	Positive MTB	Test and Control should be detected		30.37	27.98
2	Positive MTB	Test and Control should be detected		29.60	27.21
3	Positive MTB	Test and Control should be detected		30.18	27.21

^{*}Ct- Cycle Threshold

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

> Name: Dr. Praveen Kumar MK Department: Quality Control Designation: QC Head

^{*}IC- Internal Control *ND- Not Detected *NA- Not Applicable