



MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 26/06/2023

Doc ID: COA/TP/VUCPK262R

Product Name	Trueprep [®] AUTO v2 Universal Cartridge Based Sample Prep Kit
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep [®] AUTO v2
Lot No.	VUCPK262R
Mfg. Date	2023-02
Exp. Date	2025-01
Pack Size	05T, 50T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep [®] AUTO v2 Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml Graduated)	Present	Present
4	Package Insert	Present	Present

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO



Performance on Truelab® Real Time micro PCR Analyzer						
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negative Control						
1	Negative MTB	ND	Control should be detected	ND	NA	28.41
2	Negative HBV	ND		ND	NA	27.47
3	Negative HCV	ND		ND	NA	28.11
4	Negative HIV	ND		ND	NA	26.99
Positive Control						
1	Positive MTB	Test and Control should be detected		31.44	NA	28.70
2	Positive HCV	Test and Control should be detected		29.13	NA	28.57
Non Template Control Test for COVID-19						
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

*Ct-Cycle Threshold *IC- Internal Control *ND- Not Detected *NA- Not Applicable *NTC- Non Template Control

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.



Name: Dr. Praveen Kumar MK
Department: Quality Control
Designation: QC Head

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA
 Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com
 Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com
 CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO