

MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 30/06/2023

Doc ID:COA/TP/003MSPT

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack			
Generic Name	Sample pre-treament pack for MTB samples			
Lot No.	No. 003MSPT			
Mfg. Date	2023-06			
Exp. Date	2025-05			
Pack Size	05T,20T,25T, 50T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

FM/QT/224-01

Page 1 of 2



	Performan	ce on Truelab	® Real Time micro F	PCR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative (Control				
1	Negative MTB	ND	Control should be detected	ND	27.45
Positive C	ontrol				
1	Positive MTB	Test and Control should be detected		29.95	28.73
2	Positive MTB	Test and Control should be detected		29.96	28.42
Non Temp	late Control				
1	NTC	Empley of the second control of the second c	entrol should not be detected	ND	ND

^{*}Ct- Cycle Threshold

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name: Dr.Praveen Kumar MK Department: Quality Control

Por Designation: QC Head

^{*}IC- Internal Control

^{*}ND- Not Detected

^{*}NTC- Non Template Control