



MOLBIO DIAGNOSTICS PVT. LTD.
An ISO 9001 & ISO 13485 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 30/05/2020

PRODUCT :Truenat™ Universal Control Kit
LOT. NO. : UC022
MFG. DATE : 2020-04
EXP. DATE : 2021-03
PACK SIZE : 08T
STORAGE : 2°C to 40°C

A.R. No.: B-UC-022

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Negative control	Present	Present
2	Positive control (08- strip tubes containing dried down positive control)	Present	Present

Specification (Ct Value)				Observation (Ct Value)	
Sr. No.	Test	Positive control	Negative control	Positive control	Negative control
01	MTB	28.33 ± 2	ND	28.33	ND
02	MTB Plus	27.01 ± 2	ND	27.01	ND
03	Beta CoV	30.37 ± 2	ND	30.37	ND
04	MTB-RIF Dx	RIF resistance not detected	Indeterminate	RIF resistance not detected	Indeterminate
05	SARS COV-2	28.8 ± 2	ND	28.43	ND

*ND- Not detected, Ct- Cycle threshold, RIF- Rifampicin

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 E mail: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1Z0