



MOLBIO DIAGNOSTICS PVT. LTD.

An ISO 9001 & ISO 13485 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No. : 874/L

Date: 05/05/2018

PRODUCT : Truenat™ CT
LOT. NO. : CT006
MFG. DATE : 2018- 04
EXP. DATE : 2020- 03
PACK SIZE : 5T, 20T
STORAGE : 2°C to 30°C

A.R. NO. : B-CT-006

PRECAUTION: The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contain 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01no.	Pass

Performance CONTROLS			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
		CLT	IC
Specificity Tests			
Negative Control		ND	29.33
Positive Control		33.33	28.8
		32.6	27.33
		32	28.13
		31.6	29.17
Standard deviation		± 0.74	± 0.81

*ND- Not detected

*A deviation of ± 1 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE


QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1

Molbio Diagnostics Pvt. Ltd.

Office: No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA

Tel: 91 832 2458111/2458112 Email: sales@molbiodiagnostics.com

Works: Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: 91 832 6624783/6682000 Fax: 91 832 2783139 E mail: admin@molbiodiagnostics.com

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA

Website: www.molbiodiagnostics.com: CIN: U33125GA2000PTC002909