



MOLBIO DIAGNOSTICS PVT. LTD.

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000018

Date: 19/07/2020

BRAND NAME	Truenat™ SARS CoV-2
GENERIC NAME	Chip-based Real Time PCR Test for COVID-19
LOT No.	VD007
MFG. DATE	2020-07
EXP. DATE	2022-06
PACK SIZE	5T, 20T
STORAGE	2°C to 30°C
A.R. No.: B-VD-007	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.


Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 dessicant pouch and 1 filter barrier tip	5 no, 20 no.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

PERFORMANCE		
CONTROLS	(Ct Values)	
Specificity Tests	SAR	Control
Negative Control	ND	23.6
	ND	23.33
Positive Control	30.14	27.5
	29.75	28.33
	30.22	27.8
	29.25	28
	30.25	28.17
	30.43	28.4
	30.43	26.75
31.43	28.6	
Standard deviation	± 0.62	± 0.60

\*ND- Not detected, Ct- Cycle threshold, SARS- Severe Acute Respiratory Syndrome

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS QC APPROVED FOR SALE

  
QUALITY CONTROL  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

## Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-2783267 E mail: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO