



MOLBIO DIAGNOSTICS PRIVATE LIMITED
ISO 13485:2016 certified company
CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 14/07/2023
Doc ID: COA/TN/PCV003

BRAND NAME	Truenat® Positive Control Kit - Panel V		
GENERIC NAME	Control Kit for Truenat® Chip-based micro PCR Tests: H3N2, Zika, Leish, Scrub T, Cholera and Shigella		
LOT NO	PCV003		
MFG. DATE	2023-07		
EXP. DATE	2024-06		
PACK SIZE	08T		
STORAGE	2°C to 40°C	A.R. No. : B-PCV-003	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Negative control	Present	Present
2	Positive control (08- strip tubes containing dried down positive control)	Present	Present
3	Package insert	Present	Present


Observation (Ct Value)			
Sr. No.	Truenat® test name	Positive control	Negative control
1	H3N2	26.70 ± 0.26	ND
2	Zika	26.51 ± 0.23	ND
3	Leish	29.12 ± 0.68	ND
4	Scrub T	28.93 ± 0.56	ND
5	Cholera	27.92 ± 0.59	ND
6	Shigella	28.38 ± 0.60	ND

*ND- Not detected, Ct- Cycle threshold

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head

Page 1 of 1



Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO