



**MOLBIO DIAGNOSTICS PVT. LTD.**

**An ISO 9001 & ISO 13485 certified company**

**CERTIFICATE OF ANALYSIS**

Mfg. Lic. No. : 874/L

Date: 23-05-2018

PRODUCT : Truenat™ Trich  
LOT. NO. : TV006  
MFG. DATE : 2018- 04  
EXP. DATE : 2020- 03  
PACK SIZE : 05T, 20T  
STORAGE : 2°C to 30°C

A.R. NO. : B-TV-006

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	05 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01 no.	Pass

Performance			
Sr. No.	CONTROLS	Test Lot (Ct Values)	
Specificity Tests		TRV	IC
Negative Control		ND	29.57
Positive Control		31.33	31
		32.33	31.75
		33	32.17
		33.8	31
Standard deviation		± 1.00	± 0.58

\*ND- Not detected

\*A deviation of  $\pm 1$  is acceptable for Ct values

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**



**QUALITY CONTROL**  
Name: Dr. Praveen Kumar  
Designation: QC Head

Page 1 of 1

## Molbio Diagnostics Pvt. Ltd.

Regd. Office: H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA  
Works: Plot Nos. L 46, Phase II - D, Verna Industrial Estate, Verna - Goa 403 722 INDIA  
Tel: 91 832 2783267 E mail: admin@molbiodiagnostics.com  
Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com  
CIN: U33125GA2000PTC002909