



**MOLBIO DIAGNOSTICS PVT. LTD.**

**An ISO 9001 & ISO 13485 certified company**

**CERTIFICATE OF ANALYSIS**

**Mfg. Lic. No. : 874/L**

**Date: 28/11/2018**

**PRODUCT : Truenat™ Influenza A/B**  
**LOT. NO. : AB003**  
**MFG. DATE : 2018- 09**  
**EXP. DATE : 2020- 08**  
**PACK SIZE : 5T, 20T**  
**STORAGE : 2°C to 30°C**

**A.R. NO. : B-AB-003a**

**PRECAUTION:** The product is offered without warranty for any other specific use than that specified in the Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Pouch contains 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	5 nos, 20 nos.	Pass
2	Colour of desiccant pouch	Blue	Pass
3	Package insert	01no.	Pass

Performance			
CONTROLS	Test Lot (Ct Values)		
	INF A	INF B	IC
Specificity Tests			
Negative Control	ND	ND	29.14
Positive Control	29.33	ND	27.4
	ND	25.75	27.33
	27.67	25.6	27.63
	28.6	25.25	28
Standard deviation	± 0.65	± 0.24	± 0.26

\*ND- Not detected

\*A deviation of  $\leq 1.5$  is acceptable for Ct values

**MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE**



**QUALITY CONTROL**  
**Name: Dr. Praveen Kumar**  
**Designation: QC Head**

Page 1 of 1

**Molbio Diagnostics Pvt. Ltd.**

**Regd. Office:** H. No. 13, Sagar Society, Dona Paula, Panaji, N. Goa 403 004. INDIA

**Works:** Plot Nos. M 46-47, Phase III - B, Verna Industrial Estate, Verna - Goa 403 722. INDIA

**Tel:** 91 832 6624783/6682000 **Fax:** 91 832 2783139 **E mail:** admin@molbiodiagnostics.com

**Office:** No. UGO-7, Ground Floor, Ambrosia Residency, Cabesa, Alto Santacruz, Goa - 403 005. INDIA

**Tel:** 91 832 2458111/2458112 **Email:** sales@molbiodiagnostics.com

**Website:** www.molbiodiagnostics.com **CIN:** U33125GA2000PTC002909