



TruenatTM

MTB

Chip-based Real Time PCR Test for *Mycobacterium tuberculosis*

1. INTENDED USE


TruenatTM MTB (REF 601030005/601030020/601030050) is a chip-based Real Time Polymerase Chain Reaction (PCR) test for the quantitative detection and diagnosis of *Mycobacterium tuberculosis* (MTB) in human pulmonary and EPTB specimen and aids in the diagnosis of infection with MTB. **TruenatTM MTB** runs on the **Truelab[®]** Real Time micro PCR Analyzers.

2. INTRODUCTION

Tuberculosis (TB) is an infectious disease caused predominantly by the bacillus *Mycobacterium tuberculosis*. It typically affects the lungs (pulmonary TB) but can affect other sites as well (extra pulmonary TB). Tuberculosis (TB) is the second largest killer worldwide, after HIV and is the leading cause of death in HIV patients. Pulmonary TB spreads through air and is highly contagious. Over 80% of TB infections are pulmonary and if left untreated, a pulmonary TB patient can infect up to 10-15 other people through close contact over the course of a year. Due to the highly infectious nature of pulmonary TB, it is important to diagnose and treat the disease very early. Despite the availability of highly effective treatment for decades, TB remains a major global health problem mainly because of poor case detection. The most common method for diagnosing pulmonary TB worldwide is sputum smear microscopy. However sensitivity of direct smear microscopy is low and estimates range from 30% to 70%. It is even lower in case of HIV-infected patients. Culture is more sensitive than microscopy and is considered the current gold standard. Culture requires specialized and controlled laboratory facility and highly skilled manpower and takes 3 to 6 weeks to provide the result. Molecular techniques such as polymerase chain reaction (PCR) or Real Time PCR are much more sensitive than microscopy and culture. However PCR or Real Time PCR tests have so far been restricted to centralized reference laboratories as they require skilled manpower and elaborate infrastructure. Also the turnaround time for results could take a few days.

The **Truelab[®]** Real Time micro PCR System enables decentralization and near patient diagnosis of MTB by making real time PCR technology rapid, simple, robust and user friendly and offering "sample to result" capability even at resource limited settings. This is achieved through a combination of light weight, portable, mains/battery operated **Truelab[®]** Real Time micro PCR Analyzer and **Trueprep[®] AUTO/AUTO v2** Sample Prep Device and room temperature stable **TruenatTM** micro PCR chip and **Trueprep[®] AUTO/AUTO v2** Sample Prep kits so that even the peripheral laboratories with minimal infrastructure and minimally trained technicians can easily perform these tests routinely in their facilities and report PCR results in less than an hour. Moreover, with these devices PCR testing can also be initiated in the field level, on site.

TruenatTM MTB is a disposable, room temperature stable, chip-based Real Time PCR test with dried MgCl₂ in reaction well and freeze dried PCR reagents in microtube for performing Real Time PCR test for detection of *Mycobacterium tuberculosis* and runs on the **Truelab[®]** Real Time micro PCR Analyzer. It requires only six (6) µL of purified DNA to be added to the reaction well for the analysis.



The intelligent chip also carries test and batch related information including standard values for quantitation. The **TruenatTM MTB** chip-based Real Time PCR test also stores information of used test to prevent any accidental re-use of the test.

NOTE :Truelab[®] / TruenatTM / Trueprep[®] / Truepet[®] are all trademarks of Molbio Diagnostics Private Limited.

The **Truelab[®]** Real Time micro PCR Analyzer is protected by the following patents and patents granted: IN 2313/CHE/2007 (Patent No. 281573), WO2009/047804 and corresponding claims of any foreign counterpart(s) thereof.

The **TruenatTM** micro PCR chip is protected by the following patents and patents pending: IN 2312/CHE/2007, WO 2009/047805 and corresponding claims of any foreign counterpart(s) thereof. The **TruenatTM MTB** chip-based Real Time PCR test is protected by the following patents and patents pending: IN 796/CHE/2012 and corresponding claims of any foreign counterpart(s) thereof.

3. PRINCIPLE OF THE TEST

TruenatTM MTB works on the principle of Real Time Polymerase Chain Reaction based on Taqman chemistry. The DNA from the patient sample is first extracted using **Trueprep[®] AUTO/AUTO v2** Universal Cartridge Based Sample Prep Device and **Trueprep[®] AUTO/AUTO v2** Universal Cartridge Based Sample Prep Kit. The **TruenatTM MTB** chip is placed on the chip tray of the **Truelab[®]** Real Time micro PCR Analyzer. Six (6) µL of the purified DNA is then dispensed using the provided micropipette and tip into the microtube containing freeze dried PCR reagents and allowed to stand for 30-60 seconds to get a clear solution. **△ No mixing by tapping, shaking or by reverse pipetting should be done.** Six (6) µL of this clear solution is then pipetted out using the same pipette and tip and dispensed into the reaction well of the

TruenatTM MTB chip and the test is started. A positive amplification causes the dual labeled fluorescent probe in the **TruenatTM MTB** chip-based Real Time PCR test to release the fluorophores in an exponential manner which is then captured by the built-in opto-electronic sensor and displayed as amplification curve on the analyzer screen, on a real time basis during the test run. The Cycle threshold (Ct) is defined as the number of amplification cycles required for the fluorescent signal to cross the threshold (i.e. exceed the background signal). Ct levels are inversely proportional to the amount of target nucleic acid in the sample. (i.e. the lower the Ct level the greater is the amount of target nucleic acid in the sample). In the case of negative samples, amplification does not occur and a horizontal amplification curve is displayed on the screen during the test run. At the end of the test run, a MTB "DETECTED" or "NOT DETECTED" result is displayed and in positive cases, quantitative values is also displayed on the screen. Based on the Ct of the internal positive control (IPC), the validity of the test run is also displayed. The IPC is a full process control that undergoes all the processes the specimen undergoes - from extraction to amplification thereby validating the test run from sample to result. Absence of or shift of IPC Ct beyond a pre-set range in case of negative samples invalidates the test run. While IPC will co-amplify in most positive cases also, in some specimen having a high target load, the IPC may not amplify, however the test run is still considered valid. The results can be printed via Bluetooth using the **Truelab[®]** micro PCR printer or transferred to the lab computer/or any remote computer via Wifi network or 3G/GPRS network. Upto 20,000 results in **Truelab[®] Uno Dx/ Duo/Quattro** can be stored on the analyzer for future recall and reference.

4. TARGET SELECTION

The target sequence for **TruenatTM MTB** is *nrdB* gene which codes for ribonucleoside- diphosphate reductase large subunit.

5. CONTENTS OF THE TruenatTM MTB KIT

A. Individually sealed pouches, each containing

1. **TruenatTM MTB** micro PCR chip.
2. Microtube with freeze dried PCR reagents.
3. DNase & RNase free pipette tip.
4. Desiccant pouch.

B. Package Insert.

REF	601030005	601030020	601030050
△	5T	20T	50T

6. CONTENTS OF Trueprep[®] AUTO MTB Sample pre-treatment pack

- A. Liquefaction buffer
- B. Lysis buffer
- C. Disposable transfer pipette (graduated)
- D. Package Insert.

REF	60204AS05	60204AS20	60204AS50
△	5T	20T	50T

7. STORAGE AND STABILITY

TruenatTM MTB chip is stable for two (2) years from the date of manufacture if stored between 2-30°C. It is also stable for one (1) month at temperatures up to 45°C. Avoid exposure to light or elevated temperatures (above recommended levels). Do not freeze.

Trueprep[®] AUTO MTB Sample pre-treatment pack is stable for two (2) years from the date of manufacture if stored between 2-40°C. It is also stable for one (1) month at temperatures up to 45°C. Do not freeze.

8. MATERIALS REQUIRED BUT NOT PROVIDED WITH THE KIT

Truelab[®] Real Time micro PCR Workstation (REF 62301001/633010001/64301001 /653010001) consisting of

1. **Trueprep[®] AUTO / AUTO v2** Sample Prep Device (REF 603041001/ 603042001).
2. **Truelab[®] Uno Dx / Truelab[®] Duo / Truelab[®] Quattro** Real Time micro PCR Analyzer (REF603021001/603022001/603023001).
3. **Truelab[®]** micro PCR Printer (REF 603050001).
4. **Truepet[®]** SPA fixed volume precision micropipette - 6 µl (REF 604070006).
5. **Truelab[®]** Microtube Stand (REF 603070001).

Also required additionally are: **Trueprep[®] AUTO** Universal Cartridge Based Sample Prep Kit (REF 60203AR05/REF 60203AR25/REF 60203AR50) or **Trueprep[®] AUTO v2** Universal Cartridge Based Sample Prep Kit (REF 60207AR05/REF 60207AR25/60207AR50), **TruenatTM** Universal Control Kit (REF 601100008), Powder free disposable gloves, waste disposal container with lid.

9. SPECIMEN PREPARATION FOR EXTRACTION WITH Trueprep[®] AUTO/AUTO v2

TruenatTM MTB requires purified nucleic acids from pulmonary and EPTB specimen that are extracted using the **Trueprep[®] AUTO/AUTO v2** Universal Cartridge Based Sample Prep Device and **Trueprep[®] AUTO/AUTO v2** Universal Cartridge Based

Sample Prep Kit. Samples must be liquefied and pre-treated using the **Trueprep® AUTO MTB** Sample pre-treatment pack provided (Refer to the package insert of **Trueprep® AUTO MTB Sample Pre-treatment Pack** for details) before proceeding for extraction.

Sample Storage and Transportation:

Sample pre-treatment decontaminates the specimen and makes it ready for extraction. Sample in this form is stable for 3 days at upto 40°C and 1 week at 30°C.

Nucleic acid extraction: Follow Extraction procedure (section-13) of **Trueprep® AUTO/AUTO v2** Universal Cartridge Based Sample Prep Kit package insert. (Refer to the User Manual of **Trueprep® AUTO/AUTO v2** Universal Cartridge Based Sample Prep Device and the package insert of **Trueprep® AUTO** Universal Cartridge Based Sample Prep Kit for details). Dispose off lysis buffer tube and transfer pipette after use, as per the section on “Disposal and Destruction” (Section 17).

10. SAFETY PRECAUTIONS

1. For *in vitro* diagnostic use only.
2. Bring all reagents and specimen to room temperature (20 - 30°C) before use.
3. Do not use kit beyond expiry date.
4. Carefully read the User Manuals, package inserts and Material Safety Data Sheets (MSDS) of all the components of the **Truelab® Real Time micro PCR System** before use.
5. All materials of human origin should be handled as though potentially infectious.
6. Do not pipette any material by mouth.
7. Do not eat, drink, smoke, apply cosmetics or handle contact lenses in the area where testing is done.
8. Use protective clothing and wear disposable gloves when handling samples and while performing sample extraction.

11. PROCEDURAL PRECAUTIONS

1. Check all packages before using the kit. Damage to the packaging does not prevent the contents of the kit from being used. However, if the outer packaging is damaged the user must confirm that individual components of the kit are intact before using them.
2. Do not perform the test in the presence of reactive vapours (e.g. from sodium hypochlorite, acids, alkalis or aldehydes) or dust.
3. While retrieving the **Truenat™ MTB** micro PCR chip, microtube and the DNase & RNase free pipette tip from the pouch, ensure that neither bare hands nor gloves that have been used for previous tests run are used.

12. PROCEDURAL LIMITATIONS

1. Optimal performance of this test requires appropriate specimen collection, handling, storage and transport to the test site.
2. Though very rare, mutations within the highly conserved regions of the target genome where the **Truenat™** assay primers and/or probe bind may result in the under-quantitation of or a failure to detect the presence of the concerned pathogen.
3. The instruments and assay procedures are designed to minimize the risk of contamination by PCR amplification products. However, it is essential to follow good laboratory practices and ensure careful adherence to the procedures specified in this package insert for avoiding nucleic acid contamination from previous amplifications, positive controls or specimens.
4. A specimen for which the **Truenat™** assay reports “Not Detected” cannot be concluded to be negative for the concerned pathogen. As with any diagnostic test, results from the **Truenat™** assay should be interpreted in the context of other clinical and laboratory findings.

13. CLEANING AND DECONTAMINATION

1. Spills of potentially infectious material should be cleaned up immediately with absorbent paper tissue and the contaminated area should be decontaminated with disinfectants such as 0.5% freshly prepared sodium hypochlorite [10 times dilution of 5% sodium hypochlorite (household bleach)] before continuing work.
2. Sodium hypochlorite should not be used on an acid-containing spill unless the spill-area is wiped dry first. Materials used to clean spills, including gloves, should be disposed off as potentially bio-hazardous waste e.g. in a biohazard waste container.

14. TEST PROCEDURE

(Please also refer the **Truelab®** Real Time micro PCR Analyzer user manual)

1. Switch on the **Truelab®** Analyzer.
2. Select user and enter password.
3. For **Truelab® Uno Dx**, select the test profile for “MTB” to be run from the Profiles Screen on the Analyzer screen. For **Truelab® Duo/Quattro**, select the Bay (Idle1/2) for **Duo** and (Idle1/2/3/4) for **Quattro** from the Status Screen to view the Profiles Screen. Select the test profile for “MTB” to be run from the Profiles Screen, on the Analyzer screen.
4. Enter the patient details as prompted in the **Truelab®** Analyzer screen.
5. Press Start Reaction.
6. For **Truelab® Uno Dx**, Press the eject button to open the chip tray. For **Truelab® Duo/Quattro**, the chip tray opens automatically on tapping the “Start Reaction”

button.

7. Open a pouch of **Truenat™ MTB** and retrieve the micro PCR chip, microtube and DNase & RNase free pipette tip.
8. Label the chip with the patient ID using a marker pen at the space provided on the back side of the chip.
9. Place the **Truenat™ MTB** chip on the chip tray without touching the white reaction well. The reaction well should be facing up and away from the Analyzer. Gently press the chip to ensure that it is seated in the chip tray properly.
10. Place the microtube containing freeze dried PCR reagents in the microtube stand provided along with the **Truelab®** Real Time micro PCR workstation **after ensuring that white pellet of dried PCR reagents remains at the bottom of the microtube**. Remove the microtube cap and dispose it off as per the section on “Disposal and Destruction” (Section 17). Using the filter barrier tip provided in the pouch, pipette out six (6) µL of the purified DNA from the Elute Collected Tube into the microtube. Allow it to stand for 30-60 seconds to get a clear solution.
⚠ Do not mix it by tapping, shaking or by reverse pipetting. Using the same filter barrier tip, pipette out six (6) µL of this clear solution and dispense into the centre of the white reaction well of the **Truenat™ MTB** chip. Take care not to scratch the internal well surface and not to spill elute on the outside of the well. Dispose off the microtip as per the section on “Disposal and Destruction” (Section 17).
11. For **Truelab® Uno Dx**, slide the chip tray containing the **Truenat™ MTB** chip-based Real Time PCR test loaded with the sample into the **Truelab®** Analyzer. Press Done on the “Please Load Sample” Alert message. For **Truelab® Duo/Quattro**, select “YES” at the Please load Sample prompt. Chip tray will close automatically and the reaction will start.
12. Read the result from the screen.
13. After the reaction is completed, for **Truelab® Uno Dx**, push the Eject button to eject the chip tray. For **Truelab® Duo/Quattro**, tap the “Open/Close Tray” button to eject the chip tray.
14. Take out the **Truenat™ MTB** micro PCR chip at end of the test and dispose it off as per the section on “Disposal and Destruction” (Section 17).
15. Turn on **Truelab®** micro PCR printer and select print on the screen for printing out hard copy of the results. Test results are automatically stored and can be retrieved any time later. (Refer to **Truelab®** Analyzer manual).
16. Switch off the **Truelab®** Analyzer.

15. RESULTS & INTERPRETATIONS

Two amplification curves are displayed on the **Truelab®** Analyzer screen when optical plot is selected to indicate the progress of the test. Both the target and the internal positive control (IPC)* curves will take a steep, exponential path when the fluorescence crosses the threshold value in case of positive samples. The Ct will depend on the number of bacterial genomes in the sample. The target curve will remain horizontal throughout the test duration and the IPC curve will take an exponential path in case of negative samples. In case the IPC curve remains horizontal in a negative sample, the test is considered as Invalid. At the end of the test run, the results screen will display “DETECTED” for Positive result or “NOT DETECTED” for Negative result. The result screen would also display the Ct value and the colony forming units per milliliter (CFU/ml) for positive specimen. The result screen also displays the validity of the test run as “VALID” or “INVALID”. Invalid samples have to be repeated with fresh specimen from the sample preparation stage.
*Note: IPC will co-amplify in most positive cases also, in some specimen having a high target load, the IPC may not amplify, however the test run is still considered valid.

16. QUALITY CONTROL PROCEDURES

To ensure that the **Truelab®** Analyzer is working accurately, run positive and negative controls from time to time. The **Truenat™** Universal Control Kit containing Positive Control and Negative Control must be ordered separately. It is advisable to run controls under the following circumstances:

- Whenever a new shipment of test kits is received.
- When opening a new test kit lot.
- If the temperature of the storage area falls outside of 2-30° C.
- By each new user prior to performing testing on clinical specimen.

17. DISPOSAL AND DESTRUCTION

1. Submerge the used **Truenat™ MTB** chip, microtube, microtube cap, transfer pipette, pipette tips etc. in freshly prepared 0.5% sodium hypochlorite solution for 30 minutes before disposal as per the standard medical waste disposal guidelines.
2. Disinfect the solutions and/or solid waste containing biological samples before discarding them according to local regulations.
3. Samples and reagents of human and animal origin, as well as contaminated materials, disposables, neutralized acids and other waste materials must be discarded according to local regulations after decontamination by immersion in a freshly prepared 0.5% of sodium hypochlorite for 30 minutes (1 volume of 5% sodium hypochlorite for 10 volumes of water).
4. Do not autoclave materials or solutions containing sodium hypochlorite.
5. Chemicals should be handled in accordance with Good Laboratory Practice and disposed off according to the local regulations.

18. SPECIFIC PERFORMANCE CHARACTERISTICS

Analytical Exclusivity(Primer Specificity):

Genomic DNA sequence of following microorganisms were evaluated *in silico* from the the NCBI database using the NCBI nucleotide blast and primer blast tools to determine for potential cross-reactivity in the **Truenat™ MTB** assay. No cross reactivity in the performance of the **Truenat™ MTB** assay was observed with the below listed microorganisms.

Bacteria	Bacteria
<i>Escherichia coli</i>	<i>Acinetobacter anitratus</i>
<i>Enterobacter cloacae</i>	<i>Chlamydia trachomatis</i>
<i>Enterococcus faecalis</i>	<i>Gardenerella vaginalis</i>
<i>Candida albicans</i>	<i>Streptococcus mutans</i>
<i>Trichomonas vaginalis</i>	<i>Salmonella enterica</i>
<i>Staphylococcus aureus</i>	<i>Neisseria gonorrhoeae</i>
Virus	Virus
Adenovirus	Human Immunodeficiency Virus
Epstein-Barr virus	Herpes Simplex virus
Simian virus	Cytomegalovirus
Hepatitis B virus	Hepatitis C Virus


Specificity of the **Truenat™ MTB** assay was evaluated by testing Non tuberculosis mycobacterium (NTM) strains as given in the below table. No cross reactivity in the performance of the **Truenat™ MTB** assay was observed with the below listed NTM strains.

Bacteria	Bacteria
<i>M. malmoense</i>	<i>M. simiae</i>
<i>M. intracellulare</i>	<i>M. trivale</i>
<i>M. scrofulaceium</i>	<i>M. terrae</i>
<i>M. ulcerance</i>	<i>M. flavescens</i>
<i>M. abscessus</i>	<i>M. haemophilum</i>
<i>M. fortuitum</i>	<i>M. thermoresistibile</i>
<i>M. avium</i>	<i>M. marinum</i>
<i>M. gordanae</i>	<i>M. xenopi</i>
<i>M. szulgai</i>	<i>M. vaccae</i>
<i>M. kansasii</i>	<i>M. chelonae</i>
<i>M. asiaticum</i>	<i>M. smegmatis</i>
<i>M. celatum</i>	

ASSAY RANGE AND LIMIT OF DETECTION

The MTB strain H37Rv from Zeptomatrix was used for LoD determination. The LoD was determined by making dilutions of 50, 100, 250, 500 and 1000 CFU/ml of H37Rv strain in negative sputum samples. Further, the MTB DNA from each of these dilutions were extracted using **Trueprep® AUTO** Universal Cartridge Based Sample Prep Device and PCR performed on **Truelab® Uno Dx** Real Time micro PCR Analyzer using **Truenat™ MTB** chips. LoD was determined to be 100 CFU/ml sputum sample.

CFU/ml	% Positivity
50	50
100	100
250	100
500	100
1000	100



Robustness:

To determine whether the **Truenat™ MTB** chip-based Real Time PCR test showed any signs of carryover between the runs, alternate positive and negatives sputum samples were extracted and further tested the same by PCR. 20 positive samples and 20 negative samples were used for the study. The **Truenat™ MTB** test did not exhibit detectable carryover from positive to negative samples.

Reproducibility:

The purpose of this study is to compare the functional performance of the **Truenat™ MTB** assay using three different titres of samples on **Truelab® Uno Dx** real time micro PCR analyzer. High, Medium and low titre samples were extracted on **Trueprep® AUTO** sample prep device and tested among three different users(Inter user), on three different devices(Inter device) and on 5 consecutive days(Inter day) to check the variability. Mean %CV values for all titres has been calculated for Inter User (1.53), Inter day(1.74) and Inter Device (1.02) which were in the accepted range of $\leq 15\%$ CV for **Truenat™ MTB** assay.

Interfering Substances

The purpose of this study is to determine the effect of potentially interfering substances on the **Truenat™ MTB** assay. For this study low load sample has been used. To the sputum sample, 10% and 30% blood was spiked and then the sample was extracted on **Trueprep® AUTO** Universal Cartridge Based Sample Prep Device and PCR was performed on **Truelab® Uno Dx** Real Time Quantitative micro PCR analyzer using **Truenat™ MTB** assay. The presence of blood till 30% did not interfere

with the performance of **Truenat™ MTB** assay.

Precision of Truenat™ MTB assay:

Precision was tested by performing **Truenat™ MTB** assay with extracted DNA from sputum of High (2.00E+06 copies/mL), Medium (2.00E+04 copies/mL) and Low (2.00E+02 copies/mL) for five consecutive days. Every day PCR for each titre DNA was run in duplicates. The %CV values obtained for High titre (2.9), Medium titre (1.55) and low titre (2.1) were within the accepted range of $\leq 15\%$ CV for **Truenat™ MTB** assay.

Clinical Validations:

a) Clinical validation 1:

A pilot study was conducted at P. D. Hinduja National Hospital and Medical Research Centre (Nikam, Chaitali, et al. "Rapid diagnosis of *Mycobacterium tuberculosis* with **Truenat™ MTB**: a near-care approach." PLOS One 8.1 (2013): e51121), 226 sputum specimens from suspected TB patients were analyzed using smear microscopy, culture, in-house nested PCR and **Truenat™ MTB**. Pelleted sputum specimens were re-suspended in lysis buffer from the **Trueprep® MAG Sputum** kit and processed using the **Trueprep® MAG Sample Prep Device** followed by PCR on **Truenat™ MTB** chip-based Real Time PCR test. Results were compared with a Composite Reference Standard (CRS) comprising microbiological tests, clinical and radiological findings and patient history. The results are tabulated below:

	Smear		Culture		In-house Nested PCR		Truenat™ MTB	
	+ve	-ve	+ve	-ve	+ve	-ve	+ve	-ve
CRS +ve	120	71	141	50	173	18	174	17
CRS -ve	00	35	00	35	03	32	00	35
Sensitivity %	62.83		73.82		90.58		91.10	
Specificity %	100		100		91.43		100	
PPV%	100		100		98.30		100	
NPV%	33.02		41.18		64.00		67.31	

CRS - Composite Reference Standard, PPV - Positive Predictive Value, NPV - Negative Predictive Value.

The results show that **Truenat™ MTB** was the most sensitive (91.10%) and specific (100%) test compared with the Composite Reference Standard. **Truenat™ MTB** also showed high sensitivities of 99.12% among smear positive and culture positive specimen and 75.86% among smear negative and culture positive specimen.

Another study evaluating the **Truenat™ MTB** test was performed using a characterized 100 sample panel from suspected TB patients referred to a hospital in South East Asia. The study involved processing of 500µl of each sputum specimen using the **Trueprep® MAG Sputum Sample Prep Kit** on the **Trueprep® MAG Sample Prep Device**. The purified nucleic acids were tested using **Truenat™ MTB** chip-based Real Time PCR test and MTB specific primers and probe on a commercial real-time PCR machine.

Sample Type	Commercial real-time PCR machine result	Truenat™ MTB
S+C+	40/40 (100%)	40/40 (100%)
S-C+	30/40 (75%)	30/40 (75%)
S-C-	0/20 (nil detected)	0/20 (nil detected)

(S: Smear, C: Culture)

The **Truenat™ MTB** chip-based Real Time PCR test was able to detect 100% of the S+C+ samples (40/40), 75% of S-C+ samples (30/40) and gave a negative result for 100% of the S-C- samples (20/20).

b) Clinical validation 2: validation of EPTB Samples

A retrospective study was conducted at P. D. Hinduja National Hospital and Medical Research Centre, Mumbai. A total of 266 specimens from suspected EPTB patients were analyzed using culture and **Truenat™ MTB**. EPTB Specimens (Pleural/Peritoneal fluid, lymph node aspirate, Abscess, Bronchoalveolar lavage, Biopsy and Tissues) obtained from patients were processed using **Trueprep® MAG Sample Prep Device** followed by PCR on **Truelab® Uno** real time micro PCR analyzer using **Truenat™ MTB** chip-based Real Time PCR test. Sensitivity and specificity for each sample type was calculated taking Culture as 'Gold Standard'. The results are tabulated below:

Sample type/No. of Samples	BAL (n=27)	Pleural/Peritoneal Fluid (n=57)	Abscess (n=81)	Biopsy (n=49)	Lymph node (n=32)	Tissue (n=20)
Sensitivity	84.62%	44.44%	82.22%	100%	71.43%	83.33%
Specificity	71.43%	89.58%	69.44%	93.33%	77.78%	92.86%

Sensitivity and specificity for all sample types combined was calculated against culture and found to be 78.00% and 84.07% respectively. The results are tabulated below:

Overall (All Sample Types)		
	Culture +Ve	Culture -Ve
Truenat MTB +Ve	71	28
Truenat MTB -Ve	20	147

c) Clinical validation 3:

A panel of 30 samples comprising of 10 known positives and 20 known negative sputum samples were tested on three different manufacturing lots of **Truenat™ MTB** assay at National Institute for Research in Tuberculosis, Chennai against WHO approved system as comparator. DNA from 30 sputum samples were extracted using **Trueprep® AUTO** Universal Cartridge Based Sample Prep Device. The elutes were run in parallel on three lots of **Truenat™ MTB** chips.

Specificity: All 20 negative samples by comparator assay were also found to be negative by the three lots of **Truenat™ MTB** assay, showing **100%** specificity.

Sensitivity: All 10 positive sample results were correlated between the method giving a sensitivity of **100%** for the three lots of **Truenat™ MTB** assay.

Concordance: The results obtained by **Truenat™ MTB** assay on the 10 positive samples were compared with the results obtained by the comparator assay. This showed good performance and 100% concordance with the WHO approved comparator test for all three lots on the tested panel of samples. Mean Standard deviation of Ct values across the 3 lots for MTB target was 0.45 well within acceptable Ct variation of 1.66 indicates (0.5 log) showed good performance and no significant lot to lot variation.

d) Clinical validation 4:

A study evaluating the performance of **Truenat™ MTB** assay was conducted at National Reference Laboratory of Mycobacteriology, Tunisia. A total of 114 specimens (91 sputum and 23 extrapulmonary specimens) were collected from suspected TB patients were analysed. Out of 91 sputum samples, 25 sputum specimens from patients with other lung infections were taken as negative controls. All collected specimens were analysed using smear microscopy, culture, comparator PCR assay and **Truenat™ MTB** and **Truenat™ MTB-RIF** assay. Results were compared with a composite reference standard (CRS) comprising microbiological test results, clinical and radiological findings and patient history to assess the performance of the **Truenat™ MTB** assay.

Performance (% of cases detected) of molecular tests in various specimen categories:

Test	S+ (N=46)	C+ (N=49)	S+C+ (N=46)	S-C+ (N=3)
Comparator assay	100	95.9	97.8	66.6
Truenat MTB	100	100	100	100

(S: Smear, C: Culture)

Sensitivity: The **Truenat™ MTB** chip-based Real Time PCR Test was able to detect all of the S+C+ specimens (46/46) and S-C+ specimens (3/3), showing **100%** sensitivity.

Specificity: All 48 CRS negative specimens were also found to be negative by **Truenat™ MTB** assay, showing **100%** specificity.

Rifampicin Resistance: 4 strains are resistant to rifampicin by proportion method and comparator assay, 3 strains are resistant by **Truenat™ MTB-RIF** and for one strain, the test is indeterminate as quantity of specimen was not enough.

19. REFERENCES

(1) WHO Fact sheet March 2012. <http://www.who.int/mediacentre/factsheets/fs104/en/>.

(2) Todar's Online Textbook of Bacteriology - Kenneth Todar, Ph.D.

(3) WHO report 2011 Global Tuberculosis Control.

(4) Karen R Steingart et. al. (2006). Sputum processing methods to improve the sensitivity of smear microscopy for tuberculosis: a systematic review. *The Lancet Infectious Diseases*, 6(10):664 - 674.

(5) P. Farnia et. al. (2002). Improving Sensitivity of Direct Microscopy for Detection of Acid-Fast Bacilli in Sputum: Use of Chitin in Mucus Digestion, *J. Clin Microbiol.*; 40(2): 508-511.

(6) E. Ogbaini-Emovon (2009). Current Trends In The Laboratory Diagnosis of Tuberculosis, *Benin Journal of Post Graduate Medicine*, Vol.11 Supplemental, 79 - 90.

(7) Dye C, Watt CJ, Bleed DM et. al. (2005). Evolution of tuberculosis control and prospects for reducing tuberculosis incidence, prevalence and deaths globally. *JAMA* 2005; 293: 2767-2775.

(8) N. Kennedy et. al (1994). Polymerase Chain Reaction for Assessing Treatment Response in Patients with Pulmonary Tuberculosis. *The Journal of Infectious Diseases*, 170(3): 713-716.

(9) Nikam C, Jagannath M, Narayanan MM, Ramanabhiraman V, Kazi M, et al. (2013). Rapid Diagnosis of *Mycobacterium tuberculosis* with Truenat MTB: A Near-Care Approach, *PLoS ONE*, 8(1): e51121. doi:10.1371/journal.pone.0051121.

(10) Nikam C, Kaz Mi, Nair C, Jagannath M, Manoj M, et al. (2014). Evaluation of the Indian TrueNAT micro RT-PCR device with GeneXpert for case detection of pulmonary tuberculosis. *International Journal of Mycobacteriology*, 3(3):205-210, doi: org/10.1016/j.ijmyco.2014.04.003.

(11) Lee DJ, Kumarasamy N, Resch SC, Sivaramakrishnan GN, Mayer KH, Tripathy S, et al. (2019). Rapid, point-of-care diagnosis of tuberculosis with novel Truenat assay: Cost-effectiveness analysis for India's public sector. *PLoS ONE* 14(7): e0218890. <https://doi.org/10.1371/journal.pone.0218890>

SYMBOL KEYS

Consult instructions for use	In vitro Diagnostic Medical Device. Not for medicinal use.	Temperature Limitation	Catalogue Number	For single use only	Manufacturer	This Side Up
Date of Manufacture	Date of Expiry	Batch Number / Lot Number	Caution	Contains sufficient for <n> tests	Authorised Representative in the European Community	

Molbio Diagnostics Private Limited
Registered Office & Manufacturing Unit:
 Plot No. L-46, Phase II D,
 Verna Industrial Estate, Verna,
 Goa - 403 722, INDIA
www.molbiodiagnostics.com