



Truenat®

Positive Control Kit - Panel V

Control Kit for Truenat® Chip-based micro PCR Tests: H3N2, Zika, Leish, Scrub T, Cholera and Shigella

1. INTENDED USE

Truenat® Positive Control Kit - Panel V (REF 801050008) is a set of positive and negative control for validating the performance of Truenat® chip based micro PCR tests - H3N2, Zika, Leish, Scrub T, Cholera and Shigella as well as Truelab® Real Time micro PCR Analyzers.

2. INTRODUCTION

Testing for infectious diseases by detecting the pathogens nucleic acids using nucleic acid amplification methods is a highly specific and sensitive diagnostic tool. Molbio's Truelab® micro PCR System is a nucleic acid amplification platform that works on real time Polymerase Chain Reaction (PCR) technology that enables near patient diagnosis through Truenat® disposable, disease specific micro PCR chips and a portable, automated Truelab® Real Time micro PCR Analyzers.

To ensure that the Truenat® chip based micro PCR tests and the Truelab® Real Time micro PCR Analyzers are working accurately, it is necessary to run positive and negative controls from time to time. The Truenat® Positive Control kit - Panel V is a set of 1. Dried down mixture of calibrated Positive Control DNA (representing target nucleic acids of Truenat® tests as H3N2, Zika, Leish, Scrub T, Cholera and Shigella) and 2. Negative Control, that are run in place of nucleic acids extracted from specimen.

It is advisable to run controls under the following circumstances: • Whenever a new shipment of test kits is received. • When opening a new test kit lot. • If the temperature of the storage area falls outside of 2-40°C. • By each new user prior to performing testing on clinical specimen.

NOTE : Trademarks: Truelab® /Truenat® / Trueprep® / Truepet® are all trademarks of Molbio Diagnostics Private Limited.

The Truelab® Real Time micro PCR Analyzer is protected by the following patents and patents pending: IN 2313/CHE/2007, WO 2009/047804 and corresponding claims of any foreign counterpart(s) thereof.

The Truenat® micro PCR chip is protected by the following patents and patents pending: IN 2312/CHE/2007, WO 2009/047805 and corresponding claims of any foreign counterpart(s) thereof.

3. PRINCIPLE OF THE TEST

Truenat® chip based PCR tests work on the principle of Real Time Polymerase Chain Reaction based on Taqman chemistry. Six (6) µL of the reconstituted positive control or the negative control is then dispensed using the provided micropipette and tip into the microtube containing disease specific freeze dried RT PCR/PCR reagents and allowed to stand for 30-60 seconds to get a clear solution. △**No mixing by tapping, shaking or by reverse pipetting should be done.** Six (6) µL of this clear solution is then pipetted out using the same pipette and tip and dispensed into the reaction well of the disease specific Truenat® chip based micro PCR test. The Truenat® chip is then inserted in the Truelab® Real Time micro PCR Analyzer where thermal cycling takes place. A positive amplification causes the disease specific fluorescent probe in the Truenat® chip to release the fluorophores in an exponential manner which is then captured by the built-in opto-electronic sensor and displayed as amplification curve on the analyzer screen, on a real time basis during the test run. The Cycle threshold (Ct) is defined as the number of amplification cycles required for the fluorescent signal to cross the threshold (i.e. exceed the background signal). Ct value is linearly correlated with the initial load of target DNA present. The positive control in the Truenat® Positive Control Kit- Panel V contains calibrated quantities of DNA representing target nucleic acids of Truenat® tests H3N2, Zika, Leish, Scrub T, Cholera and Shigella to yield a range bound Ct value. This range is preset in the analyzer and the analyzer automatically compares this with the Ct value of the positive control. The negative control is a buffer solution which is also used to reconstitute the positive control, does not have any DNA and hence no amplification is expected to occur. In this case a horizontal amplification curve is displayed on the screen during the test run. The results screen on the analyzer displays "DETECTED" with Ct value or "NOT DETECTED" and whether the result is "Valid" or "Not valid". For a valid result, the positive control must amplify and the Ct value obtained should fall within the expected range. The negative control should not amplify. A repeatedly "Not Valid" result for positive control indicates a malfunction of either the Truenat® test or the Truelab® analyzer. Amplification ("Not Valid" result) for negative control indicates nucleic acid contamination in the testing place, equipments or reagents. An invalid result needs further investigation and resolution before carrying on any further testing. The results can be printed via Bluetooth using the Truelab® micro PCR printer or transferred to remote computer via Wifi network or 3G/GPRS network. The result is also stored in the analyzer memory for records.

4. CONTENTS OF THE Truenat® POSITIVE CONTROL KIT - PANEL V

A. Individually sealed pouch, containing

1. A strip of 8 micro tubes each containing dried down Positive Control : 1Strip.
2. A screw Cap vial containing 1 ml of Negative Control : 1 x 1 ml.
- B. Package Insert.

REF	801050008
Σ	8T

5. STORAGE AND STABILITY

Truenat® Positive Control Kit - Panel V is stable for one (1) year from the date of manufacture if stored between 2-40°C. It is also stable for upto one (1) month at temperatures up to 45°C. Avoid exposure to light or elevated temperatures (above recommended levels).

6. MATERIALS REQUIRED BUT NOT PROVIDED WITH THE KIT

Truelab® Real Time micro PCR Workstation (REF 623010001 / 633010001 / 643010001 / 653010001) consisting of

1. Truelab® Uno Dx / Truelab® Duo / Truelab® Quattro Real Time micro PCR Analyzer (REF 603021001/603022001/603023001).
2. Truelab® micro PCR Printer (REF 603050001).
3. Truepet® SPA fixed volume precision micropipette - 6 µl (REF 604070006).
4. Truelab® Microtube Stand (REF 603070001).

Also required additionally are: Truenat® Chip Based micro PCR Tests, 50 µl micropipette, Powder free disposable gloves, waste disposal container with lid.

7. CONTROL PREPARATION

The dried down positive control in the micro tube needs to be reconstituted. Open the Zip Lock pouch, retrieve the positive control strip and cut out one positive control micro tube from the strip. Replace the remaining strip into the pouch and seal with the zip lock for future use. Pipette 50 µl of negative control from the negative control vial into the micro tube using a fresh DNase/ RNase free filter barrier micropipette tip. Dispose off the tip immediately as per the section on "Disposal and Destruction" (Section 13). Mix the micro tube gently for a few minutes. The reconstituted positive control must be used immediately.

The negative control is a ready to use solution.

△ **Positive control can cause contamination and should be handled with extreme care while opening the micro tube, reconstitution and use. Avoid spillage. Dispose and destroy the micro tube with left over control and micro pipette tips as described in section 13 below.**

8. SAFETY PRECAUTIONS

1. For *in vitro* diagnostic use only.
2. Bring all reagents and specimen to room temperature (20 - 30°C) before use.
3. Do not use kit beyond expiry date.
4. Carefully read the User Manuals, package inserts and Material safety Data Sheets (MSDS) of all the components of the Truelab® Real Time micro PCR System before use.
5. Do not pipette any material by mouth.
6. Do not eat, drink, smoke, apply cosmetics or handle contact lenses in the area where testing is done.
7. Use protective clothing and wear disposable gloves all through the testing process.

9. PROCEDURAL PRECAUTIONS

1. Check all packages before using the kit. Damage to the packaging does not prevent the contents of the kit from being used. However, if the outer packaging is damaged the user must confirm that individual components of the kit are intact before using them.
2. Do not perform the test in the presence of reactive vapours (e.g. from sodium hypochlorite, acids, alkalis or aldehydes) or dust.
3. While retrieving the Truenat® chip, microtube and the DNase & RNase free pipette tip from the pouch, ensure that neither bare hands nor gloves that have been used for previous tests run are used.

10. CLEANING AND DECONTAMINATION

1. Spills of potentially infectious material should be cleaned up immediately with absorbent paper tissue and the contaminated area should be decontaminated with disinfectants such as 0.5% freshly prepared sodium hypochlorite [10 times dilution of 5% sodium hypochlorite (household bleach)] before continuing work.
2. Sodium hypochlorite should not be used on an acid-containing spill unless the spill-area is wiped dry first. Materials used to clean spills, including gloves, should be disposed off as potentially bio-hazardous waste e.g. in a biohazard waste container.

11. TEST PROCEDURE

(Please also refer the Truelab® Real Time micro PCR Analyzer user manual)

1. Switch on the Truelab® Analyzer.
2. Select user and enter password.
3. For Truelab® Uno Dx, select the test profile of the Truenat® chip based test to be run from the Profiles Screen, on the Analyzer screen. For Truelab®

Duo/Quattro, select the Bay (Idle1/2) for **Duo** and (Idle1/2/3/4) for **Quattro** from the Status Screen to view the Profiles Screen. Select the test profile of the **Truenat®** chip based test to be run from the Profiles Screen, on the Analyzer screen.

4. Select Positive Control or Negative control as the case may be from the menu under **sample type** in the **Truelab®** Real Time micro PCR Analyzer screen. Ignore all other prompts on the screen.
5. Press Start Reaction.
6. For **Truelab® Uno Dx**, Press the eject button to open the chip tray. For **Truelab® Duo/Quattro**, the chip tray opens automatically on tapping the "Start Reaction" button.
7. Open a pouch of **Truenat®** and retrieve micro PCR chip, microtube and the DNase & RNase free pipette tip the chip-based Real Time PCR test.
8. Place the **Truenat®** chip on the chip tray without touching the white reaction well. The reaction well should be facing up and away from the Analyzer. Gently press the chip to ensure that it has seated in the chip tray properly.
9. Place the microtube containing disease specific freeze dried RT PCR/PCR reagents in the microtube stand provided along with the **Truelab®** Real Time micro PCR workstation **after ensuring that white pellet of dried RT PCR/PCR reagents remains at the bottom of the microtube**. Remove the microtube cap and dispose it off as per the section on "Disposal and Destruction" (Section 13). Using the filter barrier tip provided in the pouch, pipette out Six (6) µL of the reconstituted positive control or the negative control into the microtube containing disease specific freeze dried RT PCR/PCR reagents and allowed to stand for 30-60 seconds to get a clear solution. **⚠️ No mixing by tapping, shaking or by reverse pipetting should be done**. Using the same filter barrier tip, pipette out six (6) µL of this clear solution and dispense into the centre of the white reaction well of the disease specific **Truenat®** chip based micro PCR test. Take care not to scratch the internal well surface and not to spill elute on the outside of the well. Dispose off the microtip as per the section on "Disposal and Destruction" (Section 13).
10. For **Truelab® Uno Dx**, Slide the chip tray containing the **Truenat®** chip-based Real Time PCR test loaded with the control into the **Truelab®** Analyzer. Press Done on the "Please Load Sample" Alert message. For **Truelab® Duo/Quattro**, Select "YES" at the Please load Sample prompt. Chip tray will close automatically and the reaction will start.
11. Read the result from the screen.
12. After the reaction is completed, for **Truelab® Uno Dx**, push the Eject button to eject the chip tray. For **Truelab® Duo/Quattro**, tap the "Open/Close Tray" button to eject the chip tray.
13. Take out the **Truenat®** micro PCR chip at end of the test and dispose it off as per the section on "Disposal and Destruction" (Section 13).
14. Turn on **Truelab®** micro PCR printer and select print on the screen for printing out hard copy of the results. Test results are automatically stored and can be retrieved any time later (Refer to **Truelab®** Analyzer manual).
15. Switch off the **Truelab®** Analyzer.

12. RESULTS & INTERPRETATIONS

An amplification curve is displayed on the **Truelab®** Real Time micro PCR Analyzer screen to indicate the progress of the test. The curve will take a steep, exponential path when the fluorescence crosses the threshold value in case of amplification. The curve will remain horizontal throughout the test duration in case of no amplification. At the end of the test run, the results screen will display "DETECTED" for Positive result or "NOT DETECTED" for Negative result. The result screen would also display the Ct value if amplification was detected. The result screen also displays the validity of the test run as "VALID" or "INVALID". In the case of positive control, a valid result indicates that the control amplified and the Ct value was within the expected range. For negative control, a valid result indicates that there was no amplification detected. Invalid tests have to be repeated with fresh control and if still invalid for positive control, this indicates either that the lot of **Truenat®** test is not working or that there is a problem with the **Truelab®** analyzer. Check again with a new lot of **Truenat®**. If problem persists then contact Molbio support. Repeated invalid for negative control indicates contamination of the work area, the chip tray of **Truelab®** analyzer, pipettes, or the negative control itself. Check with a fresh vial of negative control. If not corrected, decontaminate work area, the chip tray of **Truelab®** analyzer and pipette using a wipe wetted with 0.5 % sodium hypochlorite and subsequently with clean water wipe and check with negative control. Repeat this procedure till problem is resolved. If problem persists then contact Molbio support.

⚠️ Sodium hypochlorite can cause PCR reactions to fail. Ensure that the work area, surfaces and equipment are free of sodium hypochlorite or its vapor and confirm with positive control after every cleaning routine before proceeding with another test.

13. DISPOSAL AND DESTRUCTION

1. Submerge the used **Truenat®** micro PCR chip, microtube, microtube cap, pipette tips etc. in freshly prepared 0.5% sodium hypochlorite solution for 30 minutes before disposal as per the standard medical waste disposal guidelines.
2. Disinfect the solutions and/or solid waste containing biological samples before discarding them according to local regulations.
3. Samples and reagents of human and animal origin, as well as contaminated

materials, disposables, neutralized acids and other waste materials must be discarded according to local regulations after decontamination by immersion in a freshly prepared 0.5% of sodium hypochlorite for 30 minutes (1 volume of 5% sodium hypochlorite for 10 volumes of water).

4. Do not autoclave materials or solutions containing sodium hypochlorite.
5. Chemicals should be handled in accordance with Good Laboratory Practice and disposed off according to the local regulations.

14. REFERENCES

1. Shuo Y., Darren J.H., Yingfeng A. (2016) Recent Advances in Universal TA Cloning Methods for Use in Function Studies. Protein Eng Des Sel. 29 (11); 551–6.
2. Ming-Yi Z., Celso E.G. (2000) Universal TA Cloning. Curr. Issues Mol. Biol. 2(1): 1-7.
3. Hoorfar J., Malorny B., Abdulmawjood A. (2004) Practical Considerations in Design of Internal Amplification Controls for Diagnostic PCR Assays. J. Clin. Microbiol. 42(5); 1863–8.
4. Chaokun L., Aiyun W., Benchang S. (2011) FastCloning: a highly simplified, purification-free, sequence- and ligation-independent PCR cloning method. BMC Biotechnology 11(92); 1-10
5. https://tools.thermofisher.com/content/sfs/manuals/MAN0012706_InsTAClone_PCR_Cloning_UG.pdf.

SYMBOL KEYS

Molbio Diagnostics Private Limited
Registered Office & Manufacturing Unit:

Plot No. L-46, Phase II D,
Verna Industrial Estate, Verna,
Goa - 403 722, INDIA
www.molbiodiagnostics.com
Email: sales@molbiodiagnostics.com (Sales Enquiries)
customersupport@molbiodiagnostics.com (Feedback and Customer Support)

EC REP Qarad EC-REP BV, Pas 257, 2440 Geel, Belgium