

FORM MD-5

[See sub-rule (4) of rule 20 and sub-rule (6) of rule 20]

Licence to Manufacture for Sale or for Distribution of Class A or Class B medical device

Licence Number: MFG/IVD/2019/000003

Endorsement No. 1

- 1. M/s MOLBIO DIAGNOSTICS PVT. LTD, H.NO. 13, SAGAR SOCIETY, DONA PAULA, PANAJI NORTH GOA, 403 004TISWADI, South Goa, Goa (India) 403004 Telephone No.: 08322783267 FAX: 08322783139 has been licenced to manufacture for sale or for distribution the below listed medical device(s) at the premises situated at M/s Molbio Diagnostics Pvt. Ltd., Plot No. L-46, Phase II D, Verna Industrial Estate, Verna, South Goa, Goa (India) 403722 Telephone No.: 09960590604 FAX: 0832 2783267
- 2. Details of medical device(s) [Annexed]
- 3. This licence is subject to the provisions of the Medical Devices Rules, 2017 and conditions prescribed therein.

ANNEXURE

S.No.	Details Of Device(s)
1	Generic Name:Universal Cartridge Based Sample Prep Device
	Model No.:NIL
	Intended Use:Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit works with
	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Device and is used for extraction
	and purification of nucleic acids from a wide range of biological specimens. It is for In Vitro
	Diagnostics (IVD) Laboratory use only.
	Class of medical device:Class A
	Material of construction:"Instrument for Nucleic acid extraction and purification comprising
	of a 2 line alphanumeric LCD for User interface, a cartridge holder tray for holding the
	Cartridge from Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit, mechanical
	assemblies, electronic components with cable harness, in-built firmware that control all
	aspects of the system and with QR code reader. Mains / rechargeable Li-ion battery powered."
	Dimension(if any):
	Shelflife:NA
	Sterile or Non sterile:Non-Sterilized
	Brand Name(if registered under the Trade Marks Act, 1999):Trueprep® AUTO v2

Place: Directorate of Food & Drugs Administration,

Bambolim- Goa.

Date:16-Apr-20 State Licensing Authority