

MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 16/04/2025

Doc ID: COA/TP/MSPT264/50T

Product Name	Trueprep [®] AUTO MTB Sample Pre-treatment Pack
Generic Name	Sample pre-treatment pack for MTB samples
Lot No.	MSPT264
Mfg. Date	2025-04
Exp. Date	2027-03
Pack Size	50T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep[®] AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

Performance on Truelab® Real Time micro PCR Analyzer					
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative Control					
1	Negative MTB	ND	Control should be detected	ND	26.60
Positive Control					
1	Positive MTB	Test and Control should be detected		22.38	26.75
2	Positive MTB	Test and Control should be detected		30.48	28.08
Non Template Control					
1	NTC	Test and Control should not be detected		ND	ND

*Ct- Cycle Threshold *IC- Internal Control *ND- Not Detected *NTC- Non Template Control

The lot is tested and verified as per QC Standards and is approved.



Name: Dr.Praveen Kumar *Mk*
Department: Quality Control
Designation: QC Head

for