

MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 23/12/2023

Doc ID:COA/TP/035MSPT

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack			
Generic Name	Sample pre-treament pack for MTB samples			
Lot No.	035MSPT			
Mfg. Date	2023-12			
Exp. Date	2025-11			
Pack Size	25T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present



	Performan	ce on Truelat	o® Real Time micro	PCR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative	Control				herring the second seco
1	Negative MTB	ND	Control should be detected	ND	26.55
Positive (Control				
1	Positive MTB	Test and Control should be detected		26.35	26.86
2	Positive MTB	Test and Control should be detected		30.86	27.03
Non Tem	plate Control				
1	NTC	Test and Control should not be detected		ND	ND

NOS

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

> Name: Dr. Praveen Kumar MK **Department:** Quality Control Fr Designation: QC Head

^{*}NTC- Non Template Control