



MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 30/10/2024

Doc ID: COA/TP/057USPKL/50T

Product Name	Trueprep® AUTO Universal Cartridge Based Sample Prep Kit
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO
Lot No.	057USPKL
Mfg. Date	2024-09
Exp. Date	2026-08
Pack Size	50T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep® AUTO Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml Graduated)	Present	Present
4	Package Insert	Present	Present

FM/QT/224-02

Page 1 of 2

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO

Performance on Truelab® Real Time micro PCR Analyzers						
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negative Control						
1	Negative HBV	ND	Control should be detected	ND	NA	26.78
2	Negative HCV	ND		ND	NA	27.98
3	Negative HIV-1	ND		ND	NA	26.32
4	Negative MTB	ND		ND	NA	27.33
5	Negative MTB Plus	ND		ND	NA	26.28
Positive Control						
1	Positive MTB	Test and Control should be detected		30.56	NA	27.99
2	Positive HCV	Test and Control should be detected		29.29	NA	27.49
3	Positive MTB Plus	Test and Control should be detected		27.03	NA	26.96
Non Template Control Test for COVID-19						
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

*Ct-Cycle Threshold *IC- Internal Control *ND- Not Detected *NA- Not Applicable *NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.


Name: Dr.Praveen Kumar MK
Department: Quality Control
Designation: QC Head