

MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 17/08/2024

Doc ID: COA/TP/065MSPT/50T

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack		
Generic Name	Sample pre-treatment pack for MTB samples		
Lot No.	065MSPT		
Mfg. Date	2024-08		
Exp. Date	2026-07		
Pack Size	50T		
Storage condition	2°C to 40°C		

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present



	Performan	ce on Truelal	b [®] Real Time micro	PCR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative	Control				
1	Negative MTB	ND	Control should be detected	ND	27.79
Positive	Control				
1	Positive MTB	Test and Control should be detected		26.02	28.05
2	Positive MTB	Test and Control should be detected		31.79	27.56
Non Tem	plate Control				
1	NTC	SOUTH STATE OF THE	ntrol should not be etected	ND	ND

^{*}Ct- Cycle Threshold

The lot is tested and verified as per QC standards and is approved.

Name: Dr.Praveen Kumar MK
Department: Quality Control
Designation: QC Head

^{*}IC- Internal Control

^{*}ND- Not Detected

^{*}NTC- Non Template Control