



**MOLBIO DIAGNOSTICS PRIVATE LIMITED**

**EN ISO 13485:2016 certified company**

**CERTIFICATE OF ANALYSIS**

**License No.:** MFG/IVD/2021/000064

**Date:** 03/03/2024

**Doc ID:** COA/TP/092VUCPK

|                          |                                                                         |
|--------------------------|-------------------------------------------------------------------------|
| <b>Product Name</b>      | Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit             |
| <b>Generic Name</b>      | Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2 |
| <b>Lot No.</b>           | 092VUCPK                                                                |
| <b>Mfg. Date</b>         | 2024-02                                                                 |
| <b>Exp. Date</b>         | 2026-01                                                                 |
| <b>Pack Size</b>         | 50T                                                                     |
| <b>Storage condition</b> | 2°C to 40°C                                                             |

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

| <b>Sr. No.</b> | <b>Components in each kit</b>                                                               | <b>Specification</b>                  | <b>Observation</b> |
|----------------|---------------------------------------------------------------------------------------------|---------------------------------------|--------------------|
| <b>A</b>       | <b>Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)</b> | <b>(Present (P)/Not Present (NP))</b> |                    |
| 1              | Universal Reagent Pack                                                                      | Present                               | Present            |
| 2              | Universal Cartridge Pack                                                                    | Present                               | Present            |
| 3              | Disposable Transfer Pipette (3 ml Graduated)                                                | Present                               | Present            |
| 4              | Package Insert                                                                              | Present                               | Present            |



\*Ct-Cycle Threshold \*IC- Internal Control \*ND- Not Detected \*NA- Not Applicable \*NTC- Non Template Control

**Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.**



**Name:** Dr.Praveen Kumar MK  
**Department:** Quality Control  
**Designation:** QC Head