



MOLBIO DIAGNOSTICS PRIVATE LIMITED  
EN ISO 13485:2016 certified company  
CERTIFICATE OF ANALYSIS

Licence no.: MFG/IVD/2021/000071

Date: 21/08/2024

Document ID:COA/TN/105TBb/50T

|                   |                                                                     |
|-------------------|---------------------------------------------------------------------|
| Product Name      | Truenat <sup>®</sup> MTB                                            |
| Generic Name      | Chip-based Real Time PCR Test for <i>Mycobacterium tuberculosis</i> |
| Lot No.           | 105TB                                                               |
| Set No.           | b                                                                   |
| Mfg. Date         | 2024-08                                                             |
| Exp. Date         | 2026-07                                                             |
| Pack Size         | 50T                                                                 |
| Storage condition | 2°C to 30°C                                                         |

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

| Sr. No. | Components in each kit                                                             | Specification                          | Observation             |
|---------|------------------------------------------------------------------------------------|----------------------------------------|-------------------------|
| 01      | Number of pouches as per pack size                                                 | Should be present                      | Present                 |
| 02      | Pouches Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip | Should be present                      | Present                 |
| 03      | Colour of desiccant pouch                                                          | Should be present and Orange in Colour | Present / Orange Colour |
| 04      | Package Insert                                                                     | Should be present                      | Present                 |
| 05      | Seal integrity test                                                                | Should pass                            | Pass                    |

| Performance testing data               |              |              |
|----------------------------------------|--------------|--------------|
| Samples                                | Ct Values    |              |
|                                        | MTB          | Control      |
| Non Template Control                   | ND           | ND           |
| Negative Sample                        | ND           | 28.20 ± 0.35 |
| Positive sample 1                      | 30.45 ± 0.59 | 26.66 ± 0.92 |
| Positive sample 2                      | 27.99 ± 0.54 | 27.70 ± 0.23 |
| Positive Sample 3                      | 23.97 ± 0.21 | 27.33 ± 0.60 |
| ND – Not detected, Ct- Cycle threshold |              |              |

\*Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC standards and is approved.

Name : Dr. Praveen Kumar  
Department: Quality Control  
Designation: QC Head



FM/QT/219-02

Page 1 of 1

## Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. INDIA

Tel: +91-832-6724888 / 2783267 Email: admin@molbiodiagnostics.com

Email: sales@molbiodiagnostics.com Website: www.molbiodiagnostics.com

CIN: U33125GA2000PTC002909 GSTN:30AACCM8275E1ZO