



MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 21/10/2024

Doc ID: COA/TP/162VUCPKL/50T

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2
Lot No.	162VUCPKL
Mfg. Date	2024-07
Exp. Date	2026-06
Pack Size	50T
Storage condition	2°C to 40°C

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
A	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))	
1	Universal Reagent Pack	Present	Present
2	Universal Cartridge Pack	Present	Present
3	Disposable Transfer Pipette (3 ml Graduated)	Present	Present
4	Package Insert	Present	Present



Performance on Truelab® Real Time micro PCR Analyzer						
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negative Control						
1	Negative MTB	ND	Control should be detected	ND	NA	26.70
2	Negative HBV	ND		ND	NA	26.50
3	Negative HCV	ND		ND	NA	28.53
4	Negative HIV-1	ND		ND	NA	26.89
3	Negative MTB Plus	ND		ND	NA	27.05
Positive Control						
1	Positive MTB	Test and Control should be detected		31.10	NA	26.53
2	Positive HCV	Test and Control should be detected		30.52	NA	27.82
3	Positive MTB Plus	Test and Control should be detected		28.02	NA	27.06
Non Template Control Test for COVID-19						
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

*Ct-Cycle Threshold *IC- Internal Control *ND- Not Detected *NA- Not Applicable *NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.



Name: Dr. Praveen Kumar MK
Department: Quality Control
Designation: QC Head