

# Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited)
Plot No. L-46, Phase II D, Verna Industrial Estate
Verna, Goa-403722, India

🕒+91 832 6724888 🏶 www.molbiodiagnostics.com

### MOLBIO DIAGNOSTICS LIMITED

## EN ISO 13485:2016 certified company

### **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2021/000064

Date: 21/01/2025

Doc ID: COA/TP/225VUCPKL/50T

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit			
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2			
Lot No.	225VUCPKL			
Mfg. Date	2025-01			
Exp. Date	2026-12			
Pack Size	50T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation		
Α	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)	(Present (P)/Not Present (NP))			
1	Universal Reagent Pack	Present	Present		
2	Universal Cartridge Pack	Present	Present Present		
3	Disposable Transfer Pipette (3 ml Graduated)	Present	Present Present		
4	Package Insert	Present	Present		

FM/QT/224-03

Page 1 of 2



# Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) Plot No. L-46, Phase II D, Verna Industrial Estate Verna, Goa 403722, India

S+91 832 6724888 www.molbiodiagnostics.com

	Performar	ice on T	ruelab® Real Time n	nicro PCR A	nalyzer			
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)			
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)		
Negative Control								
1	Negative MTB	ND		ND	NA	27.37		
2	Negative HBV	ND		ND	NA	26.92		
3	Negative HCV	ND	Control should be detected	ND	NA	27.25		
4	Negative HIV-1	ND		ND	NA	27.40		
3	Negative MTB Plus	ND		ND	NA	27.48		
Positiv	ve Control							
1	Positive MTB	Test and Control should be detected		30.63	NA	27.09		
2	Positive HCV	Test and Control should be detected		31.62	NA	27.18		
3	Positive MTB Plus	Test and Control should be detected		28.30	NA	26.94		
Non Te	emplate Control Test	for COV	/ID-19					
1	NTC		Test 2 and Control d not be detected	ND	ND	ND		

<sup>\*</sup>Ct-Cycle Threshold \*IC- Internal Control \*ND- Not Detected \*NA- Not Applicable \*NTC- Non Template Control

The lot is tested and verified as per QC standards and is approved.

Name: Dr. Praveen Kumar M **Department:** Quality Control Designation: QC Head

FM/QT/224-03

Page 2 of 2