

MOLBIO DIAGNOSTICS PRIVATE LIMITED EN ISO 13485:2016 certified company CERTIFICATE OF ANALYSIS

Licence no.: MFG/IVD/2019/000018

Date:10/11/2023

Document ID:COA/TN/AB017a

Product Name	Truenat® Influenza A/B	
Generic Name	Chip-based Real Time Duplex PCR Test for Influenza A and Influenza B Virus	
Lot No.	AB017	
Set No.	а	
Mfg. Date	2023-11	
Exp. Date	2025-10	
Pack Size	5T, 20T, 25T	
Storage condition	2°C to 30°C	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
01	Number of pouches as per pack size	Should be present	Present
02	Pouches Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	Should be present	Present
03	Colour of desiccant pouch	Should be present and Orange in Colour	Present / Orange Colour
04	Package Insert	Should be present	Present
05	Seal integrity test	Should pass	Pass

Performance testing data					
		Ct Values			
Samples	Influenza A	Influenza B	Control		
Non Template Control	ND	ND	ND		
Negative Sample	ND	. ND	29.96 <u>+</u> 0.57		
Positive Sample 1	30.20 <u>+</u> 0.78	ND	29.63 ± 0.08		
Positive Sample 2	ND	27.67 <u>+</u> 0.59	27.34 <u>+</u> 0.57		
Positive Sample 3	29.43 + 0.46	30.98 + 0.91	30.48 + 0.41		

 $\mbox{ND}-\mbox{Not}$ detected, Ct- Cycle threshold, Mfg- Manufacturing, Exp- Expiry, GMP- Good Manufacturing Practices

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name : Dr. Praveen Rumar Department: Quality Control

Designation: QC Head

Page 1 of 1

FM/QT/219-01

Molbio Diagnostics Private Limited