

## MOLBIO DIAGNOSTICS PRIVATE LIMITED EN ISO 13485:2016 certified company CERTIFICATE OF ANALYSIS

Licence No.: MFG/IVD/2019/000018

Date:20/02/2024

Document ID:COA/TN/HS004a

Booding in 18.00% in the incidence of th		
Product Name	Truenat® HSV 1/2	
Generic Name	Chip-based Real Time PCR Test for Herpes Simplex Virus 1 and Herpes Simplex Virus 2	
Lot No.	HS004	
Set No.	a	
Mfg. Date	2024-02	
Exp. Date	2026-01	
Pack Size	5T, 20T, 25T,50T	
Storage condition	2°C to 30°C	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
01	Number of pouches as per pack size	Should be present	Present
02	Pouches Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	Should be present	Present
03	Colour of desiccant pouch	Should be present and Orange in Colour	Present / Orange Colour
04	Package Insert	Should be present	Present
05	Seal integrity test	Should pass	Pass

Performance testing data				
	Ct Values			
Samples	HSV 1/2	Control		
Non Template Control	ND	ND		
Negative Sample	ND	28.71 <u>+</u> 0.25		
Positive sample 1	21.44 ± 0.40	27.19 <u>+</u> 0.18		
Positive sample 2	25.18 <u>+</u> 0.22	26.43 <u>+</u> 0.15		
Positive Sample 3	29.22 <u>+</u> 0.97	27.49 ± 0.30		
ND - Not detected, Ct- Cycle threshold				

\*Mfg- Manufacturing, Exp- Expiry, GMP- Good Manufacturing Practices

Manufacturing records are verified, checked and product is manufactured by the GMP, samples tested and lot is QC approved for sale purpose.

Name : Dr. Prayeen Kumar Department: Quality Control Designation: QC Head

FM/QT/219-01 Molbio Diagnostics Private Limited age 1 of 1