

## MOLBIO DIAGNOSTICS PRIVATE LIMITED EN ISO 13485:2016 certified company CERTIFICATE OF ANALYSIS

Licence No.: MFG/IVD/2019/000018

Date: 24/08/2024

Document ID: COA/TN/IN008a/25T

Product Name	Truenat® MTB - INH			
Generic Name	Chip-based Real Time PCR Test for Isoniazid Resistant Mycobacterium tuberculosis			
Lot No.	IN008			
Set No.	a			
Mfg. Date	2024-07			
Exp. Date	2026-06			
Pack Size	25T			
Storage condition	2°C to 30°C			

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
01	Number of pouches as per pack size	Should be present	Present
02	Pouches Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip	Should be present	Present
03	Colour of desiccant pouch	Should be present and Orange in Colour	Present / Orange Colour
04	Package Insert	Should be present	Present
05	Seal integrity test	Should pass	Pass

Performance testing data							
	Tm Values			Results			
Samples	KG	INH	Control				
Positive sample 1	72.98± 0.46	65.79 ± 0.33	71.31 ± 0.41	INH Resistance detected			
Positive sample 2	73.52 ± 0.63	70.52 ± 0.14	71.33 ± 0.41	INH Resistance not detected			
Positive sample 3	73.27 ± 0.69	70.57 ± 0.38	71.60 ± 0.30	INH Resistance not detected			
Tm- Melting temprate	ure, Ct- Cycle th	reshold					

\*Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC standards and is approved.

Name : Dr. Praveen Kumar Department: Quality Control

Designation: QC Head

Molbio Diagnostics Private Limited

Regd. Office & Works: Plot No. L-46, Phase II D, Verna Industrial Estate, Verna - Goa 403 722. WDIA Page 1 of 1