

MOLBIO DIAGNOSTICS PRIVATE LIMITED

ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date:03/01/2024

Doc ID:COA/TP/MSPT190

Product Name	ct Name Trueprep® AUTO MTB Sample Pre-treatment Pack			
Generic Name	Sample pre-treament pack for MTB samples			
Lot No.	MSPT190			
Mfg. Date	2024-01			
Exp. Date	2025-12			
Pack Size	5T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation		
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not	(Present (P)/Not Present (NP))		
1	Lysis Buffer	Present	Present		
2	Liquefaction Buffer	Present Present			
3	Disposable Transfer Pipette (1 ml Graduated)	Present Present			
4	Package Insert	Present	Present		

FM/QT/224-01

Page 1 of 2



	Performan	ce on Tru	elab [®] Real Time mici	o PCR Ana	ılyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
1	Specificity Tests	Test	IC	Test 1	Test 2	IC
			Negative Control			
1	Negative MTB	ND	Control should be detected	ND	NA	28.58
			Positive Control	1		
1	Positive MTB	Test and Control should be detected		29.63	NA	28.80
2	Positive MTB	Test and Control should be detected		25.17	NA	28.91

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name: Dr.Prave Department: O Designation: C