

MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 19/07/2024

Doc ID:COA/TP/MSPT226/50T

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack			
Generic Name	Sample pre-treament pack for MTB samples			
Lot No.	MSPT226			
Mfg. Date	2024-07			
Exp. Date	2026-06			
Pack Size	50T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Components in each kit	Specification	Observation
Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
Lysis Buffer	Present	Present
Liquefaction Buffer	Present	Present
Disposable Transfer Pipette (1 ml Graduated)	Present	Present
Package Insert	Present	Present
	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack) Lysis Buffer Liquefaction Buffer Disposable Transfer Pipette (1 ml Graduated)	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack) Lysis Buffer Present Liquefaction Buffer Disposable Transfer Pipette (1 ml Graduated) Present



:	Performan	ce on Truelab	o [®] Real Time micro	PCR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative	Control				
1	Negative MTB	ND	Control should be detected	ND	27.36
Positive	Control				
1	Positive MTB	Test and Control should be detected		30.62	26.80
2	Positive MTB	Test and Control should be detected		25.82	27.00
Non Tem	plate Control				
. 1	NTC		ntrol should not be etected	ND	ND

^{*}Ct- Cycle Threshold

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

> Name: Dr.Praveen Kumar MK **Department: Quality Control**

for Designation: QC Head

^{*}IC- Internal Control *ND- Not Detected

^{*}NTC- Non Template Control