

**MOLBIO DIAGNOSTICS PRIVATE LIMITED**  
**EN ISO 13485:2016 certified company**  
**CERTIFICATE OF ANALYSIS**

Licence No.: MFG/IVD/2019/000018

Date: 13/01/2025

Document ID: COA/TN/TB277/50T

|                   |                                                                     |
|-------------------|---------------------------------------------------------------------|
| Product Name      | Truenat <sup>®</sup> MTB                                            |
| Generic Name      | Chip-based Real Time PCR Test for <i>Mycobacterium tuberculosis</i> |
| Lot No.           | TB277                                                               |
| Set No.           | NA                                                                  |
| Mfg. Date         | 2025-01                                                             |
| Exp. Date         | 2026-12                                                             |
| Pack Size         | 50T                                                                 |
| Storage condition | 2°C to 30°C                                                         |

**PRECAUTION:** Refer Safety Data Sheet and Instructions for Use.

| Sr. No. | Components in each kit                                                             | Specification                          | Observation             |
|---------|------------------------------------------------------------------------------------|----------------------------------------|-------------------------|
| 01      | Number of pouches as per pack size                                                 | Should be present                      | Present                 |
| 02      | Pouches Contents: 1 chip, 1 micro tube, 1 desiccant pouch and 1 filter barrier tip | Should be present                      | Present                 |
| 03      | Colour of desiccant pouch                                                          | Should be present and Orange in Colour | Present / Orange Colour |
| 04      | Package Insert                                                                     | Should be present                      | Present                 |
| 05      | Seal integrity test                                                                | Should pass                            | Pass                    |

| Performance testing data |              |              |
|--------------------------|--------------|--------------|
| Samples                  | Ct Values    |              |
|                          | MTB          | Control      |
| Non Template Control     | ND           | ND           |
| Negative Sample          | ND           | 27.83 ± 0.51 |
| Positive sample 1        | 22.53 ± 0.37 | 26.98 ± 0.18 |
| Positive sample 2        | 27.45 ± 0.11 | 26.96 ± 0.84 |
| Positive Sample 3        | 29.52 ± 0.59 | 25.72 ± 0.30 |

ND – Not detected, Ct- Cycle threshold

\*Mfg- Manufacturing, Exp- Expiry

The lot is tested and verified as per QC Standards and is approved.



Name : Dr. Praveen Kumar  
Department: Quality Control  
Designation: QC Head

*for*