

# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India U+91832 6724888 www.molbiodiagnostics.com

#### **MOLBIO DIAGNOSTICS LIMITED**

### EN ISO 13485:2016 certified company

#### **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 13/06/2025

Doc ID:COA/TP/VUCPK436L/50T

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit				
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2				
Lot No.	VUCPK436L				
Mfg. Date	2025-04				
Exp. Date	2027-03				
Pack Size	50T				
Storage condition	2°C to 40°C				

**PRECAUTION:** Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation					
A	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)							
1	Universal Reagent Pack	Should be Present	Present					
2	Universal Cartridge Pack	Should be Present	Present					
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	Present					
4	Package Insert	Should be Present	Present					
5	desiccant pouch	Should be Present	Present					
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour					
В	Seal integrity test	Should Pass	Pass					

FM/QT/224-03

Page 1 of 2



## **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+918326724888 www.molbiodiagnostics.com

	Performand	e on Tru	elab <sup>®</sup> Real Time	micro PCR	Analyzer	· · · · · · · · · · · · · · · · · · ·
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negat	ive Control		J	307	(**************************************	(recorde)
1	Negative MTB	ND		ND	NA	27.32
2	Negative HBV	ND	Control should be detected	ND	NA	26.68
3	Negative HCV	ND		ND	NA	27.14
4	Negative HIV-1	ND		ND	NA	26.76
5	Negative MTB Plus	ND		ND	NA	27.73
Positiv	ve Control					
1	Positive MTB	Test and Control should be detected		31.07	NA	26.90
2	Positive HCV	Test and Control should be detected		31.56	NA	26.90
3	Positive MTB Plus	Test and Control should be detected		30.95	NA	28.56
Non Te	emplate Control Test	for COVI	D-19			<u> </u>
1	NTC	Contro	1, Test 2 and I should not be detected	ND	ND	ND

\*Ct-Cycle Threshold Template Control

\*IC- Internal Control

\*ND- Not Detected

\*NA- Not Applicable

\*NTC- Non

The lot is tested and verified as per QC Standards and is approved.

Name: Dr.Praveen Kumar MK **Department:** Quality Control

Designation:QC Head

FM/QT/224-03

Page 2 of 2