

MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

Mfg. Lic. No.: MFG/IVD/2019/000003

Date: 11/07/2025

Doc ID: COA/TN/PCI046

BRAND NAME	Truenat [®] Positive Control Kit - Panel I
GENERIC NAME	Control Kit for Truenat [®] Chip-based micro PCR Tests: MTB, MTB Plus, MTB-RIF Dx, Beta CoV, SARS CoV-2, COVID-19 and H1N1
LOT NO	PCI046
MFG. DATE	2025-07
EXP. DATE	2026-06
PACK SIZE	08T
STORAGE	2°C to 40°C
A.R. No. : B-PCI-046	

PRECAUTION: Refer Safety Data Sheet and Instructions for Use.

Sr. No.	Components in each kit	Specification	Observation
1	Negative control	Present	Present
2	Positive control (08- strip tubes containing dried down positive control)	Present	Present
3	Package insert	Present	Present

Observation (Ct Value)			
Sr. No.	Truenat [®] test name	Positive control	Negative control
1	MTB	28.19 ± 0.36	ND
2	MTB Plus	26.74 ± 0.41	ND
3	MTB-RIF Dx	RIF Resistance not detected	NA
4	SARS CoV-2	26.62 ± 0.38	ND
5	COVID-19	Orf 1A: 27.57 ± 0.53 E gene: 27.08 ± 0.14	ND
6	H1N1	27.38 ± 0.54	ND

*ND- Not detected, Ct- Cycle threshold, RIF- Rifampicin, NA- Not Applicable

*A deviation of ≤ 1.5 is acceptable for Ct values

MANUFACTURING RECORDS ARE VERIFIED, CHECKED AND PRODUCT IS MANUFACTURED AS PER THE MASTER FORMULA, SAMPLES TESTED AND LOT IS RELEASED FOR SALE

for 
QUALITY CONTROL
Name: Dr. Praveen Kumar
Designation: QC Head