

Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+91832 6724888 www.molbiodiagnostics.com

MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 15/07/2025

Doc ID:COA/TP/VUCPK443L/50T

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit				
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2				
Lot No.	VUCPK443L				
Mfg. Date	2025-06				
Exp. Date	2027-05				
Pack Size	50T				
Storage condition	2°C to 40°C				

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit Specification		Observation				
А	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)						
1	Universal Reagent Pack Should be Present		Present				
2	Universal Cartridge Pack	Should be Present	Present				
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	Present				
4	Package Insert	Should be Present	Present				
5	desiccant pouch	Should be Present	Present				
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour				
В	Seal integrity test	Should Pass	Pass				

FM/QT/224-03

Page 1 of 2



Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India

C+918326724888	⊕ www.molbiodia	agnostics.com
----------------	-----------------	---------------

	Performano	e on Tru	elab® Real Time	micro PCR A	nalyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negat	ive Control					
1	Negative MTB	ND		ND	NA	27.09
2	Negative HBV	ND	Control should be detected	ND	NA	27.59
3	Negative HCV	ND		ND	NA	26.55
4	Negative HIV-1	ND		ND	NA	27.53
5	Negative MTB Plus	ND		ND	NA	27.74
Positi	ve Control					
1	Positive MTB	Positive MTB Test and Control should be detected		30.70	NA	27.31
2	Positive HCV	Test and Control should be detected		31.96	NA	28.47
3	Positive MTB Plus	Test and Control should be detected		30.77	NA	28.35
Non T	emplate Control Test	for COV	ID-19			
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

*Ct-Cycle Threshold Template Control

*IC- Internal Control

*ND- Not Detected

*NA- Not Applicable

*NTC-Non

The lot is tested and verified as per QC Standards and is approved.

for Name: Dr. Praveen Kumar MK **Department:** Quality Control Designation: QC Head

FM/QT/224-03

Page 2 of 2