

Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India

MOLBIO DIAGNOSTICS LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 23/07/2025

Doc ID:COA/TP/VUCPK448L/50T

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit				
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2				
Lot No.	VUCPK448L				
Mfg. Date	2025-07				
Exp. Date	2027-06				
Pack Size	50T				
Storage condition	2°C to 40°C				

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation					
Α	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)							
1	Universal Reagent Pack	Should be Present	Present					
2	Universal Cartridge Pack	niversal Cartridge Pack Should be Present						
3	Disposable Transfer Pipette (3 ml Graduated)	Should be Present	Present					
4	Package Insert	Should be Present	Present					
5	desiccant pouch	desiccant pouch Should be Present						
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour					
В	Seal integrity test	Should Pass	Pass					

FM/QT/224-03

Page 1 of 2



Molbio Diagnostics Limited

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+918326724888 www.molbiodiagnostics.com

	Performand	e on Tru	elab® Real Time	micro PCR A	nalyzer					
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)					
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)				
Negati	Negative Control									
1	Negative MTB	ND		ND	NA	27.77				
2	Negative HBV	ND	Control should be detected	ND	NA	27.44				
3	Negative HCV	ND		ND	NA	27.37				
4	Negative HIV-1	ND		ND	NA	26.31				
5	Negative MTB Plus	ND		ND	NA	28.27				
Positiv	ve Control									
1	Positive MTB	Test and Control should be detected		30.67	NA	26.53				
2	Positive HCV	Test and Control should be detected		30.34	NA	28.19				
3	Positive MTB Plus	Test and Control should be detected		30.21	NA	27.16				
Non Te	emplate Control Test	for COVI	D-19		territorio de la competitución					
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND				

*Ct-Cycle Threshold *IC- Internal Control *ND- Not Detected *NA- Not Applicable *NTC- Non Template Control

The lot is tested and verified as per QC Standards and is approved.

Name: Dr.Praveen Kumar MK **Department:** Quality Control Designation: QC Head

FM/QT/224-03

Page 2 of 2