

# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India S+918326724888 www.molbiodiagnostics.com

### **MOLBIO DIAGNOSTICS LIMITED**

## EN ISO 13485:2016 certified company

### **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 31/07/2025

Doc ID:COA/TP/VUCPK455L/50T

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit			
Generic Name Sample preparation kit for nucleic acid extraction on Trueprep® A				
Lot No.	VUCPK455L			
Mfg. Date	2025-07			
Exp. Date	2027-06			
Pack Size	50T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit Specification		Observation				
Α	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)						
1	Universal Reagent Pack	versal Reagent Pack Should be Present					
2	Universal Cartridge Pack	Universal Cartridge Pack Should be Present					
3	Disposable Transfer Pipette (3 ml Graduated)						
4	Package Insert	Should be Present	Present				
5	Desiccant pouch Should be		Present				
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour				
В	Seal integrity test	Seal integrity test Should Pass					

FM/QT/224-03 Page 1 of 2



# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa-403722, India 

	Performano	e on Tru	elab® Real Time	micro PCR A	nalyzer					
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)					
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)				
Negati	Negative Control									
1	Negative MTB	ND		ND	NA	26.31				
2	Negative HBV	ND	Control should be detected	ND	NA	27.80				
3	Negative HCV	ND		ND	NA	26.53				
4	Negative HIV-1	ND		ND	NA	27.21				
5	Negative MTB Plus	ND		ND	NA	28.56				
Positiv	ve Control									
1	Positive MTB	Test and Control should be detected		30.29	NA	26.38				
2	Positive HCV	Test and Control should be detected		30.79	NA	27.91				
3	Positive MTB Plus	Test and Control should be detected		30.14	NA	28.34				
Non To	emplate Control Test	for COV	ID-19	<u> </u>						
1	NTC	Contro	1, Test 2 and I should not be detected	ND	ND	ND				

\*Ct-Cycle Threshold \*IC- Internal Control \*ND- Not Detected \*NA- Not Applicable \*NTC- Non Template Control

The lot is tested and verified as per QC Standards and is approved.

Name: Dr. Praveen Kumar MK **Department:** Quality Control Designation:QC Head

Page 2 of 2

FM/QT/224-03