

# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa - 403722, India

S+918326724888 www.molbiodiagnostics.com

#### **MOLBIO DIAGNOSTICS LIMITED**

### EN ISO 13485:2016 certified company

### **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 02/09/2025

Doc ID:COA/TP/VUCPK473L

Product Name	Trueprep® AUTO v2 Universal Cartridge Based Sample Prep Kit			
Generic Name	Sample preparation kit for nucleic acid extraction on Trueprep® AUTO v2			
Lot No.	VUCPK473L			
Mfg. Date	2025-07			
Exp. Date	2027-06			
Pack Size	50T			
Storage condition	2°C to 40°C			

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation					
A	Contents in the carton (Trueprep® AUTO v2 Universal Cartridge Based Sample prep kit)							
1	Universal Reagent Pack	Should be Present	Present					
2	Universal Cartridge Pack	Should be Present	Present					
3	Disposable Transfer Pipette (3 ml Graduated) Should be Present		Present					
4	Package Insert	Should be Present	Present					
5	Desiccant pouch	Desiccant pouch Should be Present						
6	Colour of desiccant pouch	Should be Orange in Colour	Orange Colour					
В	Seal integrity test	Seal integrity test Should Pass						

FM/QT/224-03

Page 1 of 2



# **Molbio Diagnostics Limited**

(Formerly Molbio Diagnostics Private Limited) L-46, Phase II D, Verna Industrial Estate Verna, Goa - 403722, India

S+91 832 6724888 www.molbiodiagnostics.com

	Performand	e on Tru	elab <sup>®</sup> Real Time	micro PCR A	nalyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Values)	
	Specificity Tests	Test	IC	Test 1 (Average)	Test 2 (Average)	IC (Average)
Negati	ve Control					
1	Negative MTB	ND		ND	NA	26.60
2	Negative HBV	ND	Control should be detected	ND	NA	27.97
3	Negative HCV	ND		ND	NA	28.50
4	Negative HIV-1	ND		ND	NA	27.57
5	Negative MTB Plus	ND		ND	NA	27.77
Positiv	ve Control					
1	Positive MTB	Test and Control should be detected		30.82	NA	26.80
2	Positive HIV-1	Test and Control should be detected		31.46	NA	26.67
3	Positive MTB Plus	Test and Control should be detected		32.83	NA	27.86
Non Te	emplate Control Test	for COVI	D-19			
1	NTC	Test 1, Test 2 and Control should not be detected		ND	ND	ND

\*Ct-Cycle Threshold \*IC- Internal Control \*ND- Not Detected \*NA- Not Applicable \*NTC- Non Template Control

The lot is tested and verified as per QC Standards and is approved.

Jou Name: Dr. Praveen Kumar MK **Department:** Quality Control **Designation: QC Head** 

FM/QT/224-03

Page 2 of 2